

TruLab Albumin U/CSF Level 1

Assayed quality control material for monitoring assay performance of quantitative in vitro determination of albumin

Order Information

5 9710 99 10 046 Level 1 3 x 1 mL

Description

TruLab Albumin U/CSF is an aqueous control containing human blood material (plasma). The albumin concentration in Level 1 is in the physiological range.

Storage

The vials both opened and unopened must be stored at 2 – 8°C.

Stability

Unopened: Until the end of the indicated month of expiry
Opened: At least 30 days

Proper storage and handling of this product must be observed.

Warnings and Precautions

1. Each individual blood donation used for production of TruLab Albumin U/CSF was found to be non-reactive when tested with approved methods for HBsAg, anti-HIV 1 + 2 and anti-HCV. As there is no possibility to exclude definitely that products derived from human blood transmit infectious agents, it is recommended to handle the control with the same precautions used for patient specimens.
2. Contains sodium azide (0.95 g/L) as preservative. Do not swallow! Avoid contact with skin and mucous membranes.
3. TruLab Albumin U/CSF contains animal material. The controls should be handled as potentially infectious and with the same precautions used for patient specimens.
4. Please refer to the safety data sheets and take the necessary precautions for the use of calibrators and controls.
5. For professional use only!

Preparation

TruLab Albumin U/CSF is liquid and ready to use.

Procedure

Please refer to the reagent package insert for instructions for use.

Assay Values

TruLab Albumin U/CSF assay values have been determined by calibration with reference preparation of the IRMM (Institute for Reference Materials and Measurements) ERM®-DA470k/IFCC in accordance with established protocols.

Assay values may vary slightly with different reagents and methodologies used. The assay values listed below are specific for this lot number of control only.

Literature

1. Röhle G, Siekmann L. Quality assurance of quantitative determination. In: Thomas L, editor. Clinical laboratory diagnostics. 1st ed. Frankfurt: TH-Books Verlagsgesellschaft; 1998. p. 1393-1401.
2. Dati F. Reference materials and guidelines for standardization of methods in laboratory medicine. In: Thomas L, editor. Clinical laboratory diagnostics. 1st ed. Frankfurt: TH-Books Verlagsgesellschaft; 1998. p. 1402-1426.

Waste management

Please refer to local legal requirements.

Manufacturer

DiaSys Diagnostic Systems GmbH
Alte Strasse 9
65558 Holzheim Germany



	Lot No.	Expiry date	Assay value	Range
TruLab Albumin U/CSF Level 1	31206	2022-11-19	25.4 mg/L	18.3 – 32.5 mg/L