

HbA1c Hemolyzing Solution

Order Information

Cat. No. 1 4570 99 10 113 Kit size 1 x 500 mL

Intended Use

Hemolyzing solution for quantitative in vitro determination of hemoglobin A1c (HbA1c) in whole blood on automated photometric systems.

Description

One HbA1c Hemolyzing Solution is used for sample preparation of whole blood, calibrators, and controls for determination of HbA1c with DiaSys test one HbA1c FS (Cat. No: 1 3329..).

Reagent

Components and Concentrations

Especially treated water

Storage and Stability

The hemolyzing solution is stable up to the date of expiry indicated on the kit, if stored at 2 - 25°C and contamination is avoided. Do not freeze and protect from light.

Warnings and Precautions

1. Please refer to the safety data sheets and take the necessary precautions for the use of laboratory reagents. For diagnostic purposes, the results should always be assessed with the patient's medical history, clinical examinations and other findings.
2. For professional use only.

Waste Management

Refer to local legal requirements.

Reagent Preparation

The reagent is ready to use.

Sample Preparation

Hemolyzing Solution	1000 µL
Sample/Calibrator/Control	20 µL

Mix and allow to stand for 5 minutes or until complete lysis is evident.



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim
Germany
www.diasys-diagnostics.com