



## Bibliografía

1. Guder WG, Zawta B et al. The Quality of Diagnostic Samples. 1<sup>st</sup> ed. Darmstadt: GIT Verlag; 2001; p. 52-3; 54-5.
2. Young DS. Effects of Drugs on Clinical Laboratory Tests. 5th. ed. Volume 1 and 2. Washington, DC: The American Association for Clinical Chemistry Press, 2000.
3. Felgenhauer K. Laboratory diagnosis of neurological diseases. In: Thomas L. Clinical Laboratory Diagnostics. 1<sup>st</sup> ed. Frankfurt: TH-Books Verlagsgesellschaft; 1998. p. 1308-26.
4. Boege F. Urinary proteins. In: Thomas L. Clinical Laboratory Diagnostics. 1<sup>st</sup> ed. Frankfurt: TH-Books Verlagsgesellschaft; 1998. p. 382-400.
5. Orsonneau JL, Douet P, Massoubre C, Lustenberger P, Bernard S. An improved pyrogallol red-molybdate method for determining total urinary protein. *Clin Chem* 1989; 35: 2233-6.
6. Watanabe N, Kamei S, Ohkubo A, Yamanaka M, Ohsawa S, Makino K et al. Urinary protein as measured with a pyrogallol red-molybdate complex manually and in a Hitachi 726 automated analyzer. *Clin Chem* 1986; 32: 1551-4.
7. Johnson AM, Rohlfs EM, Silverman LM. Proteins. In: Burtis CA, Ashwood ER, editors. *Tietz Textbook of Clinical Chemistry*. 3<sup>rd</sup> ed. Philadelphia: W.B Saunders Company; 1999. p. 477-540.
8. Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: mechanisms, detection and prevention. *Clin Chem Lab Med* 2007; 45(9):1240-1243.



## Fabricante

DiaSys Diagnostic Systems GmbH  
Alte Strasse 9 65558 Holzheim Alemania

# Total protein UC FS

## Application for urine samples

This application was set up and evaluated by DiaSys. It is based on the standard equipment at that time and does not apply to any equipment modifications undertaken by unqualified personnel.

Identification	
This method is usable for analysis:	Yes
Twin reaction:	No
Name:	TPU
Shortcut:	
Reagent barcode reference:	051
Host reference:	051

Results	
Decimals	1
Units	mg/L
Correlation factor-Offset	0.0000
Correlation factor-Slope	1.0000

Technic	
Type:	End point
First reagent:[µL]	200
Blank reagent	Yes
Sensitive to light	
Second reagent:[µL]	
Blank reagent	
Sensitive to light	
Main wavelength:[nm]	600
Secondary wavelength:[nm]	700
Polychromatic factor:	1.0000
1 st reading time [min:sec]	(-00:12)
Last reading time [min:sec]	10:00
Reaction way:	Increasing
Linear Kinetics	
Substrate depletion: Absorbance limit	
Linearity: Maximum deviation [%]	
Fixed Time Kinetics	
Substrate depletion: Absorbance limit	
Endpoint	
Stability: Largest remaining slope	
Prozone Limit [%]	

Range	
Gender	All
Age	
SERUM	>=24.0 <=141.0 mg/24h
URINE	
PLASMA	
CSF	< 500.0 mg/L
Whole blood	
Gender	
Age	
SERUM	
URINE	
PLASMA	
CSF	
Whole blood	

Reagents	
Decimals	
Units	

Contaminants	
Please refer to r910 Carryover Pair Table	
Calibrators details	
Calibrator list	Concentration
Cal. 1/Blank	0
Cal. 2	*
Cal. 3	
Cal. 4	
Cal. 5	
Cal. 6	
Max delta abs.	
Cal. 1	0.01
Cal. 2	0.02
Cal. 3	
Cal. 4	
Cal. 5	
Cal. 6	
Drift limit [%]	0.80

Sample	
Diluent	DIL A (NaCl)
Hemolysis:	
Agent [µL]	0 (no hemolysis)
Cleaner	
Sample [µL]	0
Technical limits	
Concentration technical limits-Lower	35.0000
Concentration technical limits-Upper	2900.0000
SERUM	
Normal volume [µL]	6.0
Normal dilution (factor)	1
Below normal volume [µL]	
Below normal dilution (factor)	
Above normal volume [µL]	6.0
Above normal dilution (factor)	6
URINE	
Normal volume [µL]	6.0
Normal dilution (factor)	1
Below normal volume [µL]	
Below normal dilution (factor)	
Above normal volume [µL]	6.0
Above normal dilution (factor)	6
PLASMA	
Normal volume [µL]	6.0
Normal dilution (factor)	1
Below normal volume [µL]	
Below normal dilution (factor)	
Above normal volume [µL]	6.0
Above normal dilution (factor)	6
CSF	
Normal volume [µL]	6.0
Normal dilution (factor)	1
Below normal volume [µL]	
Below normal dilution (factor)	
Above normal volume [µL]	6.0
Above normal dilution (factor)	6
Whole blood	
Normal volume [µL]	6.0
Normal dilution (factor)	1
Below normal volume [µL]	
Below normal dilution (factor)	
Above normal volume [µL]	6.0
Above normal dilution (factor)	6

Calculations	
Model	X
Degree	1

\* Enter calibrator value