

# GLDH FS\* (DGKC)

## Order Information

Cat. No.	Kit size
1 2411 99 10 021	R1 5 x 20 mL + R2 1 x 25 mL
1 2411 99 10 930	R1 4 x 20 mL + R2 2 x 10 mL

Kits for use in conjunction with DiaSys CE applications.

## Intended Use

Diagnostic reagent for quantitative in vitro determination of glutamate dehydrogenase (GLDH) activity in human serum or heparin plasma on automated photometric systems.

## Summary

Glutamate dehydrogenase (GLDH) is a mitochondrial enzyme, which is localized in all tissues but predominantly expressed in the liver. The main function of GLDH is to catalyze the nitrogen clearance from the organism. Significant elevations of the GLDH activity are detected in necrosis of hepatocytes, in acute toxic liver necrosis and in hypoxic liver diseases. Moreover, measurement of GLDH activity is a pivotal tool for the evaluation of the damage severity of parenchymal cells and for the indication of alcohol addiction [1]. In conjunction with the transaminases ALAT/GPT and ASAT/GOT, GLDH assessment is mainly utilized for differential diagnosis of liver disorders. The calculation of the (ALAT+ASAT)/GLDH ratio enables to differentiate between inflammatory liver diseases and liver diseases in which necrosis is the predominant event [1,2].

## Method

Optimized UV test, according to recommendations of the DGKC (German Society of Clinical Chemistry) [3]



One unit of GLDH is the amount of enzyme that will convert 1.0  $\mu\text{mol}$  of  $\alpha$ -ketoglutarate, NADH and  $\text{NH}_4^+$  to L-glutamate,  $\text{NAD}^+$  and  $\text{H}_2\text{O}$  per minute at the enzyme specific conditions.

## Reagents

### Components and Concentrations

R1:	Triethanolamine	pH 8.0	75 mmol/L
	$\alpha$ -Ketoglutarate		10 mmol/L
	Ammonium acetate		150 mmol/L
	EDTA		3.75 mmol/L
	ADP		1.5 mmol/L
	LDH		$\geq 2.3 \text{ kU/L}$
R2:	NADH		1.3 mmol/L

## Storage and Stability

Reagents are stable up to the date of expiry indicated on the kit, if stored at 2 – 8°C and contamination is avoided. Do not freeze and protect from light.

The open-vial stability of the reagent is 18 months until expiry date.

## Warnings and Precautions

- The reagents contain sodium azide (0.95 g/L) as preservative. Do not swallow! Avoid contact with skin and mucous membranes.
- Reagent 1 contains material of biological origin. Handle the product as potentially infectious according to universal precautions and good clinical laboratory practice.
- In very rare cases, samples of patients with gammopathy might give falsified results [4].
- Sulfasalazine and sulfapyridine medication may cause false results in patient samples. Blood collection must be performed prior to drug administration.
- In case of product malfunction or altered appearance that could affect the performance, contact the manufacturer.
- Any serious incident related to the product must be reported to the manufacturer and the competent authority of the Member State where the user and/or patient is located.
- Please refer to the safety data sheets (SDS) and take the necessary precautions for the use of laboratory reagents. For diagnostic purposes, the results should always be assessed

with the patient's medical history, clinical examinations and other findings.

- For professional use only.

## Waste Management

Refer to local legal requirements for chemical disposal regulations as stated in the relevant SDS to determine the safe disposal.

Warning: Handle waste as potentially biohazardous material. Dispose of waste according to accepted laboratory instructions and procedures.

## Reagent Preparation

The reagents are ready to use.

## Materials Required

General laboratory equipment

## Specimen

Human serum or heparin plasma

Only use suitable tubes or collection containers for specimen collection and preparation.

When using primary tubes, follow the manufacturer's instructions.

Stability [5]:

7 days	at	20 – 25°C
7 days	at	4 – 8°C
4 weeks	at	-20°C

Only freeze once. Discard contaminated specimens.

## Assay Procedure

### Basic settings for respons<sup>®</sup>940

Wavelength	340/376 nm
Temperature	37°C
Measurement	Kinetic
Sample/Calibrator	24 $\mu\text{L}$
Reagent 1	160 $\mu\text{L}$
Reagent 2	40 $\mu\text{L}$
Addition reagent 2	Cycle 25 (225 s)
Absorbance	Cycle 30/50 (270 s/450 s)
Calibration	Linear

## Calculation

### With Calibrator

$$\text{GLDH [U/L]} = \frac{\Delta A/\text{min Sample}}{\Delta A/\text{min Cal.}} \times \text{Conc. Cal. [U/L]}$$

### Conversion Factor

$$\text{GLDH [U/L]} \times 0.0167 = \text{GLDH [\mukat/L]}$$

## Calibrators and Controls

DiaSys TruCal U is recommended for calibration. Calibrator values have been made traceable to the molar extinction coefficient. Use DiaSys TruLab N and P for internal quality control. All target values of the controls are traceable to DiaSys reagent/calibrator system. Quality control must be performed after calibration. Control intervals and limits have to be adapted to the individual requirements of each laboratory. Results must be within the defined ranges. Follow the relevant legal requirements and guidelines. Each laboratory should establish corrective action in case of deviations in control recovery.

	Cat. No.	Kit size
TruCal U	5 9100 99 10 063	20 x 3 mL
	5 9100 99 10 064	6 x 3 mL
TruLab N	5 9000 99 10 062	20 x 5 mL
	5 9000 99 10 061	6 x 5 mL
TruLab P	5 9050 99 10 062	20 x 5 mL
	5 9050 99 10 061	6 x 5 mL

## Performance Characteristics

### Data evaluated on respons<sup>®</sup>940

Measuring range from 3 U/L up to 130 U/L. Linearity < 10 U/L is given with ± 3 U/L, between 10 U/L to 30 U/L within ± 10%, at > 30 U/L within ± 5%.

When values exceed this range, samples should be diluted 1 + 5 with NaCl solution (9 g/L) and the result multiplied by 6.

Limit of detection**	3 U/L
Limit of quantitation**	3 U/L

Interference by	Interferences ≤ 10% up to	Analyte concentration [U/L]
<b>Ascorbic acid</b>	70 mg/dL	14.2
	70 mg/dL	23.5
<b>Bilirubin</b> (conjugated)	6 mg/dL	7.78
	60 mg/dL	22.0
<b>Bilirubin</b> (unconjugated)	13 mg/dL	6.39
	20 mg/dL	18.1
<b>Hemolysis</b>	119 mg/dL	6.84
	300 mg/dL	21.0
<b>Lipemia</b>	200 mg/dL	7.71
	200 mg/dL	22.0

For further information on interfering substances, refer to the literature [6-8].

Precision			
Repeatability (n=20)	Sample 1	Sample 2	Sample 3
Mean [U/L]	8.37	15.7	40.3
CV [%]	2.26	2.18	2.00
Within-laboratory (n=80)	Sample 1	Sample 2	Sample 3
Mean [U/L]	7.84	15.2	41.1
CV [%]	5.91	3.36	3.97

Method comparison (n=155)	
Test x	Competitor GLDH (cobas c 501)
Test y	DiaSys GLDH FS DGKC (respons <sup>®</sup> 940)
Slope	0.965
Intercept	0.022 U/L
Coefficient of correlation	0.993

\*\* according to CLSI document EP17-A2, Vol. 32, No. 8

## Reference Range [1]

	U/L	µkat/L
<b>Children</b>		
0 – 30 day(s)	< 10	< 0.167
1 – 6 month(s)	< 7	< 0.117
7 – 12 months	< 6	< 0.100
13 – 24 months	< 5	< 0.084
2 – 3 years	< 4	< 0.067
4 – 6 years	< 5	< 0.084
<b>Adults</b>		
Female	< 5	< 0.084
Male	< 7	< 0.117

Each laboratory should check if the reference ranges are transferable to its own patient population and determine own reference ranges if necessary.

## Literature

1. Thomas L. Clinical Laboratory Diagnostics [Internet]. Prof. Lothar Thomas; 2024 [cited 2024 June 12]. Available from: <https://www.clinical-laboratory-diagnostics.com/>
2. Moss DW, Henderson AR. Clinical enzymology. In: Burtis CA, Ashwood ER, editors. Tietz Textbook of Clinical Chemistry. 3rd ed. Philadelphia: W.B Saunders Company; 1999. p. 617–721.
3. Recommendations of the German Society for Clinical Chemistry. Standardisation of methods for the estimation of enzyme activities in biological fluids. Experimental basis for the optimized standard conditions. Z Klin Chem Klin Biochem. 1972;10:182-92.
4. Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: Mechanism, detection and prevention. Clin Chem Lab Med. 2007; 45:1240-1243.
5. Guder WG, da Fonseca-Wollheim F, Heil W, Schmitt Y, Töpfer G, Wisser H, Zawta B. Quality of Diagnostic Samples. 3rd edition; 2010. p. 46-7.
6. Young DS. Effects of Drugs on Clinical Laboratory Tests. 5th ed. Volume 1 and 2. Washington DC: The American Association for Clinical Chemistry Press; 2000.
7. Young DS. Effects on Clinical Laboratory Tests - Drugs Disease, Herbs & Natural Products [Internet]. AACC Press and John Wiley and Sons, Inc; 2021 [cited 2022 February]. Available from: <https://clinfx.wiley.com/aaccweb/aacc/>
8. Sonntag O, Scholer A. Drug interference in clinical chemistry: recommendation of drugs and their concentrations to be used in drug interference studies. Ann Clin Biochem. 2001;38:376-85.

Additions and/or changes in the document are highlighted in grey. Deletions are communicated via customer info by stating the edition no. of the package insert/instruction for use.



DiaSys Diagnostic Systems GmbH  
Alte Strasse 9 65558 Holzheim  
Germany  
[www.diasys-diagnostics.com](http://www.diasys-diagnostics.com)

\* Fluid Stable