

UIBC FS*

Order Information

Cat. No. 1 1921 99 10 930 Kit size R1 4 x 20 mL + R2 2 x 10 mL

Kit for use in conjunction with DiaSys CE applications.

Intended Use

Diagnostic reagent for quantitative in vitro determination of UIBC (unsaturated iron binding capacity) in human serum or heparin plasma on automated photometric systems.

Summary

The measurement of unsaturated iron binding capacity (UIBC) in combination with serum iron is a useful diagnostic tool in the determination of various iron disorders [1]. The sum of UIBC and serum iron gives a value for the total iron binding capacity (TIBC). TIBC represents the maximum concentration of iron that serum proteins can bind. Serum UIBC levels vary in disorders of iron metabolism where iron capacities are often increased in iron deficiency and decreased in chronic inflammatory disorders or malignancies [2].

Method

Photometric test using Ferene

A known ferrous ion concentration incubated with sample, binds specifically with transferrin at unsaturated iron binding sites. Remaining unbound ferrous ions are measured with the ferene reaction. The difference between the amount of excess iron and the total amount added to the serum is equivalent to the quantity bound to transferrin. This is the UIBC of the sample.



Reagents

Components and Concentrations

R1: Buffer	pH 8.7	100 mmol/L
Ammonium iron (II) sulfate		13 µmol/L
Thiourea		120 mmol/L
R2: Ascorbic acid		240 mmol/L
Ferene		6 mmol/L
Thiourea		125 mmol/L

Storage and Stability

Reagents are stable up to the date of expiry indicated on the kit, if stored at 2 – 8°C and contamination is avoided. Do not freeze and protect from light.

The open-vial stability of the reagent is 18 months until expiry date.

Warnings and Precautions

- Components contained in UIBC FS are classified according to EC regulation 1272/2008 (CLP) as follows:



⚠ Reagent 1: Danger. Contains Dodecan-1-ol, ethoxylated and Alcohols, C9-11-iso-, C10-rich, ethoxylated. H318 Causes serious eye damage. P280 Wear protective gloves/protective clothing/eye protection. P305+P351+P338 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. P310 Immediately call a POISON CENTER/doctor.

- Reagent 1 contains sodium azide (0.95 g/L) as preservative. Do not swallow! Avoid contact with skin and mucous membranes.
- In very rare cases, samples of patients with gammopathy might give falsified results [3].
- In case of product malfunction or altered appearance that could affect the performance, contact the manufacturer.
- Any serious incident related to the product must be reported to the manufacturer and the competent authority of the Member State where the user and/or patient is located.
- Please refer to the safety data sheets (SDS) and take the necessary precautions for the use of laboratory reagents. For diagnostic purposes, the results should always be assessed

with the patient's medical history, clinical examinations and other findings.

- For professional use only.

Waste Management

Refer to local legal requirements for chemical disposal regulations as stated in the relevant SDS to determine the safe disposal.

Warning: Handle waste as potentially biohazardous material. Dispose of waste according to accepted laboratory instructions and procedures.

Reagent Preparation

The reagents are ready to use.

Materials Required

General laboratory equipment

Specimen

Human serum or heparin plasma

Only use suitable tubes or collection containers for specimen collection and preparation.

When using primary tubes, follow the manufacturer's instructions.

Separate serum/plasma at the latest 2 h after blood collection to minimize hemolysis.

Stability in serum [4]:

5 days	at	20 – 25°C
1 month	at	2 – 8°C
1 month	at	-20°C

Stability in plasma [4]:

1 month	at	2 – 8°C
1 month	at	-20°C

Only freeze once. Discard contaminated specimens.

Assay Procedure

Basic settings for BioMajesty® JCA-BM6010/C

Wavelength	596/694 nm
Temperature	37°C
Measurement	Endpoint
Sample/Calibrator	6.0 µL
Reagent 1	80 µL
Reagent 2	20 µL
Addition reagent 2	Cycle 19 (286 s)
Absorbance 1	Cycle 17/18 (231 s/244 s)
Absorbance 2	Cycle 41/42 (586 s/600 s)
Calibration	Linear

Calculation

With Calibrator

$$\text{UIBC} [\mu\text{g/dL}] = \frac{\Delta A \text{ Sample}}{\Delta A \text{ Cal.}} \times \text{Conc. Cal.} [\mu\text{g/dL}]$$

Conversion Factor

$$\text{UIBC} [\mu\text{g/dL}] \times 0.1791 = \text{UIBC} [\mu\text{mol/L}]$$

$$\text{TIBC} [\mu\text{g/dL}] = \text{UIBC} [\mu\text{g/dL}] + \text{Iron} [\mu\text{g/dL}]$$

$$\text{Transferrin} [\text{mg/dL}] = 0.7 \times \text{TIBC} [\mu\text{g/dL}]$$

Calibrators and Controls

DiaSys TruCal U is recommended for calibration. Calibrator values have been made traceable to a measurement of transferrin and iron. The transferrin value is traceable to ERM®-DA470k/IFCC and the iron value is traceable to NIST SRM 682. Use DiaSys TruLab N for internal quality control. All target values of the controls are traceable to DiaSys reagent/calibrator system. Quality control must be performed after calibration. Control intervals and limits have to be adapted to the individual requirements of each laboratory. Results must be within the defined ranges. Follow the relevant legal requirements and guidelines. Each laboratory should establish corrective action in case of deviations in control recovery.

	Cat. No.	Kit size	
TruCal U	5 9100 99 10 063	20	x 3 mL
	5 9100 99 10 064	6	x 3 mL
TruLab N	5 9000 99 10 062	20	x 5 mL
	5 9000 99 10 061	6	x 5 mL

Performance Characteristics

Data evaluated on BioMajesty® JCA-BM6010/C

Measuring range up to 625 µg/dL, linearity is given within ± 5%.
When values exceed this range, samples should be diluted 1 + 2 with NaCl solution (9 g/L) and the result multiplied by 3.

Limit of detection**	14 µg/dL
----------------------	----------

Interference by	Interferences ≤ 10% up to	Analyte concentration [µg/dL]
Ascorbic acid	30 mg/dL	304
Bilirubin (conjugated)	60 mg/dL	303
Bilirubin (unconjugated)	60 mg/dL	303
Hemolysis	200 mg/dL	303
Lipemia (triglycerides)	2000 mg/dL	293

For further information on interfering substances, refer to the literature [5,6].

Precision			
Repeatability (n=20)	Sample 1	Sample 2	Sample 3
Mean [µg/dL]	141	232	421
CV [%]	1.07	0.809	1.42
Between day (n=20)	Sample 1	Sample 2	Sample 3
Mean [µg/dL]	143	231	422
CV [%]	2.08	1.05	1.36

Method comparison (n=129)	
Test x	DiaSys UIBC FS (Hitachi 917)
Test y	DiaSys UIBC FS (BioMajesty® JCA-BM6010/C)
Slope	1.01
Intercept	-6.14 µg/dL
Coefficient of correlation	0.999

** lowest measurable concentration which can be distinguished from zero; mean + 3 SD (n = 20) of an analyte free specimen.

Reference Range

Taking into account reference values for iron and transferrin the following reference range results for UIBC [7,8]:

120 – 470 µg/dL 21 – 84 µmol/L

Each laboratory should check if the reference ranges are transferable to its own patient population and determine own reference ranges if necessary.

Literature

1. Fairbanks VF, Klee GG. Biochemical aspects of hematology. In: Burtis CA, Ashwood ER, editors. Tietz Textbook of Clinical Chemistry. 3rd ed. Philadelphia: W.B Saunders Company; 1999. p. 1642–1710.
2. Wick et al, Clinical Aspects and Lab Iron Metabolism, Anemias. Novel concepts, Springer, 5th ed. Wien New York 2003 p. 141–147.
3. Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: mechanisms, detection and prevention. Clin Chem Lab Med 2007; 45(9):1240-1243.
4. Data on file at DiaSys Diagnostic Systems GmbH.
5. Young DS. Effects of Drugs on Clinical Laboratory Tests. 5th ed. Volume 1 and 2. Washington, DC: The American Association for Clinical Chemistry Press 2000.
6. Young DS. Effects on Clinical Laboratory Tests - Drugs Disease, Herbs & Natural Products, <https://clinfx.wiley.com/aaccweb/aacc/>, accessed in June 2024. Published by AACC Press and John Wiley and Sons, Inc.
7. Dati F, Schumann G, Thomas L, Aguzzi F, Baudner S, Bienvenu J et al. Consensus of a group of professional societies and diagnostic companies on guidelines for interim reference ranges for 14 proteins in serum based on the standardization against the IFCC/BCR/CAP reference material (CRM 470). Eur J Clin Chem Clin Biochem 1996; 34: 517-20.
8. Thomas L. Clinical Laboratory Diagnostics. 1st ed. Frankfurt: TH-Books Verlagsgesellschaft; 1998. p. 273-5.

Additions and/or changes in the document are highlighted in grey. Deletions are communicated via customer info by stating the edition no. of the package insert/instruction for use.



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim
Germany
www.diasys-diagnostics.com

* Fluid Stable