

# TruLab CRP hs

**Assayed quality control material for monitoring assay performance of quantitative in vitro determination of C-reactive protein (CRP)**

## Order information

5 9730 99 10 046 Level 1 3 x 1 mL

## Description

TruLab CRP hs is a liquid stable control based on human blood material (serum) and a specific buffer. It was developed as a control for high sensitive CRP tests.

## Storage

The vials of TruLab CRP hs both opened and unopened must be stored at 2 – 8°C.

## Stability

Unopened: until the end of the indicated month of expiry

Opened: at least 3 months

Proper storage and handling of this product must be observed.

## Warnings and Precautions

1. Each individual blood donation used for production of TruLab CRP hs was found to be non-reactive when tested with approved methods for HBsAg, anti-HIV 1+2 and anti-HCV. As there is no possibility to exclude definitely that products derived from human blood transmit infectious agents, it is recommended to handle the control with the same precautions used for patient specimens.
2. Contains sodium azide (0.95 g/L) as preservative. Do not swallow! Avoid contact with skin and mucous membranes.
3. Please refer to the safety data sheets and take the necessary precautions for the use of calibrators and controls.
4. For professional use only!

## Preparation

TruLab CRP hs is liquid and ready to use.

## Procedure

Please refer to the reagent package insert for instructions for use.

## Assay Value


The assay value has been determined with DiaSys CRP U-hs calibrated by TruCal CRP hs. Assay values may vary slightly with different reagents and methodologies used. The assay values listed below are specific for this lot number of control only.

## Literature

1. Röhle G, Siekmann L. Quality assurance of quantitative determination. In: Thomas L, editor. Clinical laboratory diagnostics. 1<sup>st</sup> ed. Frankfurt: TH-Books Verlagsgesellschaft; 1998. p. 1393-1401.
2. Biosafety in Microbiological and Biomedical Laboratories. U.S. Department of Health and Human Services, Washington 1993 (HHS Publication No. [CDC] 93-8395).
3. Dati F. Reference materials and guidelines for standardization of methods in laboratory medicine. In: Thomas L, editor. Clinical laboratory diagnostics. 1<sup>st</sup> ed. Frankfurt: TH-Books Verlagsgesellschaft; 1998. p. 1393-1401.

## Manufacturer

DiaSys Diagnostic Systems GmbH  
Alte Strasse 9  
65558 Holzheim Germany



	Lot No.	Expiry date	Assay Value	Range
TruLab CRP hs Level 1	38020	2027-05-16	0.994 mg/L	0.726 – 1.26 mg/L