

Total protein FS*

Order Information

Cat. No. 1 2311 99 10 962
 Kit size Σ 1890 (R1: 6 x 315, R2: 6 x 315)

Intended Use

Diagnostic reagent for quantitative in vitro determination of total protein in human serum or heparin plasma on automated BioMajesty® JCA-BM6010/C.

Summary

Measurement of total protein is a useful test in a variety of disorders. Decreased total protein concentrations can be detected in defective protein synthesis in the liver, protein loss due to impaired kidney function, intestinal malabsorption or nutritional deficiency. Elevated protein levels occur in chronic inflammatory disorders, liver cirrhosis and dehydration. [1,2]

Method

Photometric test according to biuret method

Proteins form a violet blue color complex with copper ions in alkaline solution. The absorbance of the color is directly proportional to the concentration.

Reagents

Components and Concentrations

R1:	Sodium hydroxide	100 mmol/L
	Potassium sodium tartrate	17 mmol/L
R2:	Sodium hydroxide	500 mmol/L
	Potassium sodium tartrate	80 mmol/L
	Potassium iodide	75 mmol/L
	Copper sulphate	30 mmol/L

Storage and Stability

The reagents are stable up to the date of expiry indicated on the kit, if stored at 2 - 25°C and contamination is avoided. Protect reagents from light.

Warnings and Precautions

- The reagents contain sodium hydroxide. Do not swallow! Avoid contact with skin and mucous membranes.
- Δ Reagent 1: Warning. H290 May be corrosive to metals. P234 Keep only in original packaging. P390 Absorb spillage to prevent material damage.
- Δ Reagent 2: Warning. Contains: Potassium iodide. H290 May be corrosive to metals. H315 Causes skin irritation. H319 Causes serious eye irritation. H373 May cause damage to organs through prolonged or repeated exposure. H412 Harmful to aquatic life with long lasting effects. P234 Keep only in original packaging. P273 Avoid release to the environment. P280 Wear protective gloves/protective clothing/eye protection. P305+P351+P338 If in eyes: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. P314 Get medical advice/attention if you feel unwell.
- In serum or plasma of patients who have received large intravenous amounts of polydextrans, too high values can be measured with the biuret method. In such cases an alternative method (e.g. Kjeldahl) has to be used.
- In very rare cases, samples of patients with gammopathy might give falsified results [3].
- Please refer to the safety data sheets and take the necessary precautions for the use of laboratory reagents. For diagnostic purposes, the results should always be assessed with the patient's medical history, clinical examinations and other findings.
- For professional use only.

Waste Management

Refer to local legal requirements.

Reagent Preparation

The reagents are ready to use. The bottles are placed directly into the reagent rotor.

Materials Required

General laboratory equipment

Specimen

Human serum or heparin plasma

Stability [4]:

6 days	at	20 – 25°C
4 weeks	at	4 – 8°C
at least one year	at	-20°C

Only freeze once. Discard contaminated specimens.

Calibrators and Controls

DiaSys TruCal U is recommended for calibration. TruCal U calibrator values have been made traceable to the biuret method. Use DiaSys TruLab N and P for internal quality control. Each laboratory should establish corrective action in case of deviations in control recovery.

	Cat. No.	Kit size
TruCal U	5 9100 99 10 063	20 x 3 mL
	5 9100 99 10 064	6 x 3 mL
TruLab N	5 9000 99 10 062	20 x 5 mL
	5 9000 99 10 061	6 x 5 mL
TruLab P	5 9050 99 10 062	20 x 5 mL
	5 9050 99 10 061	6 x 5 mL

Performance Characteristics

Exemplary data mentioned below may slightly differ in case of deviating measurement conditions.

Measuring range up to 14 g/dL. In case of higher concentrations re-measure samples after manual dilution with NaCl solution (9 g/L) or use rerun function.	
Limit of detection**	0.05 g/dL
Onboard stability	7 days
Calibration stability	7 days
Interfering substance	Interferences $\leq 10\%$ up to
Ascorbic acid	30 mg/dL
Bilirubin (conjugated and unconjugated)	60 mg/dL
Hemoglobin	500 mg/dL
Lipemia (triglycerides)	1000 mg/dL
For further information on interfering substances refer to Young DS [5,6].	

Precision			
Within run (n=20)	Sample 1	Sample 2	Sample 3
Mean [g/dL]	4.78	6.17	7.40
CV [%]	0.57	0.52	0.35
Between day (n=20)	Sample 1	Sample 2	Sample 3
Mean [g/dL]	5.97	6.63	7.13
CV [%]	1.00	1.00	1.15

Method comparison (n=100)	
Test x	Competitor Total protein
Test y	DiaSys Total protein FS
Slope	1.00
Intercept	0.040 g/dL
Coefficient of correlation	0.998

** lowest measurable concentration which can be distinguished from zero; mean + 3 SD (n = 20) of an analyte free specimen.

Reference Range [1]

	[g/dL]	
Adults	6.6 – 8.8	
Children	Female	Male
1 – 30 day(s)	4.2 – 6.2	4.1 – 6.3
1 – 6 month(s)	4.4 – 6.6	4.7 – 6.7
6 months – 1 year	5.6 – 7.9	5.5 – 7.0
1 – 18 year(s)	5.7 – 8.0	5.7 – 8.0

Each laboratory should check if the reference ranges are transferable to its own patient population and determine own reference ranges if necessary.

Literature

1. Thomas L. Clinical Laboratory Diagnostics. 1st ed. Frankfurt: TH-Books Verlagsgesellschaft; 1998. p. 644-7.
2. Johnson Am, Rohlf's EM, Silverman LM. Proteins. In: Burtis CA, Ashwood ER, editors. Tietz Textbook of Clinical Chemistry. 3rd ed. Philadelphia: W.B Saunders Company; 1999. p. 477-540.
3. Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: mechanisms, detection and prevention. ClinChemLabMed 2007;45(9):1240-1243.
4. Guder WG, Zawta B et al. The Quality of Diagnostic Samples. 1st ed. Darmstadt: GIT Verlag; 2001; p. 42-3.
5. Young DS. Effects of Drugs on Clinical Laboratory Tests. 5th ed. Volume 1 and 2. Washington, DC: The American Association for Clinical Chemistry Press 2000.
6. Young DS. Effects on Clinical Laboratory Tests - Drugs Disease, Herbs & Natural Products, <https://clinf.wiley.com/aaccweb/aacc/>, accessed in April 2021. Published by AACC Press and John Wiley and Sons, Inc.



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

* Fluid Stable

Total protein FS

Chemistry code 10 231

Application for serum and plasma samples

This application was set up and evaluated by DiaSys. It is based on the standard equipment at that time and does not apply to any equipment modifications undertaken by unqualified personnel.

Analytical Conditions	
R1 volume	80
R2e volume	0
R2 volume	20
R1 diluent vol	0
R2e diluent vol	0
R2 diluent vol	0
Sample vol (S)	2
Sample vol (U)	2
Reagent 1 mix	weak
Reagent 2e mix	weak
Reagent 2 mix	weak
Reaction time	10

Sub-analy. Conditions	
Name	TP
Digits	2
M-wave L.	545
S-wave.L	****
Analy.mthd.	EPA
Calc.mthd.	STD
Qualit. judge	No

Analysis Test Condition Setting (M)		
Sample Type	Serum	Urine
Reac. sample vol.	2	2
Diluent method	No dil	No dil
Undil. sample vol.	0	0
Diluent volume	0	0
Diluent position	0	0

entered by user

Endpoint method	
Re.absorb (u)	9.999
Re. Absorb (d)	-9.999

Calculation Method Setting	
M-DET.P.l	0
M-DET.P.m	41
M-DET.P.n	42
S-DET.P.p	17
S-DET.P.r	18
Check D.P.l.	0
Limit value	0.003
Variance	10
Reac.type	Inc

Reaction Rate Method	
Cycle	2
Factor	2
E2 corre	Not do
Blank (u)	9.999
Blank (d)	-9.999
Sample (u)	9.999
Sample (d)	-9.999

Standards Setting	
FV	#
BLK H	9.999
BLK L	-9.999
STD H	9.999
STD L	-9.999