

## Iron FS\* Ferene

### Order Information

**Cat. No.**

1 1911 99 10 921

**Kit size**



480 (4 x 120)

### Intended Use

Diagnostic reagent for quantitative in vitro determination of iron in human serum or heparin plasma on automated respons<sup>®</sup>920.


### Summary

Iron exists in the body as a component of hemoglobin and myoglobin as well as bound to transferrin for the transport in plasma and stored in ferritin. Increased iron concentrations occur in hemochromatosis and liver damage [1]. Malabsorption due to gastrointestinal diseases can cause decreased iron levels, and may thus lead to anemia. Blood loss after gastrointestinal lesions or heavy menstrual bleeding can generate anemia, too [2].

### Method

Photometric test using Ferene

Iron bound to transferrin (Fe<sup>3+</sup>) is completely released under acidic conditions and reduced to Fe<sup>2+</sup>. Iron forms a blue complex with Ferene. The absorbance at 595 nm is directly proportional to the iron concentration.



### Reagents

#### Components and Concentrations

<b>R1:</b>	Acetate buffer	pH 4.5	1 mol/L
	Thiourea		120 mmol/L
<b>R2:</b>	Ascorbic acid		240 mmol/L
	Ferene		3 mmol/L
	Thiourea		120 mmol/L

### Storage and Stability

Reagents are stable up to the date of expiry indicated on the kit, if stored at 2 – 8°C and contamination is avoided. Do not freeze and protect from light.

The open-vial stability of the reagent is 18 months until expiry date.

### Warnings and Precautions

- Components contained in Iron FS Ferene are classified according to EC regulation 1272/2008 (CLP) as follows:



**⚠ Reagent 1: Danger.** Contains Dodecan-1-ol, ethoxylated and Alcohols, C9-11-iso-, C10-rich, ethoxylated. H315 Causes skin irritation. H318 Causes serious eye damage. P264 Wash hands and face thoroughly after handling. P280 Wear protective gloves/protective clothing/eye protection. P305+P351+P338 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. P310 Immediately call a POISON CENTER/doctor.

- To avoid carryover interference, please take care of efficient washing especially after use of interfering reagents. Please refer to the DiaSys respons<sup>®</sup>920 Carryover Pair Table. Carryover pairs and automated washing steps with the recommended cleaning solution can be specified in the system software. Please refer to the user manual.
- In very rare cases, samples of patients with gammopathy might give falsified results [3].
- In case of product malfunction or altered appearance that could affect the performance, contact the manufacturer.
- Any serious incident related to the product must be reported to the manufacturer and the competent authority of the Member State where the user and/or patient is located.
- Please refer to the safety data sheets (SDS) and take the necessary precautions for the use of laboratory reagents. For diagnostic purposes, the results should always be assessed with the patient's medical history, clinical examinations and other findings.
- For professional use only.

### Waste Management

Refer to local legal requirements for chemical disposal regulations as stated in the relevant SDS to determine the safe disposal.

Warning: Handle waste as potentially biohazardous material. Dispose of waste according to accepted laboratory instructions and procedures.

### Reagent Preparation

The reagents are ready to use. The bottles are placed directly into the reagent rotor.

### Materials Required

General laboratory equipment

### Specimen

Human serum or heparin plasma

Only use suitable tubes or collection containers for specimen collection and preparation.

When using primary tubes, follow the manufacturer's instructions.

Separate serum/plasma at the latest 2 h after blood collection to minimize hemolysis.

#### Stability [4]:

7 days	at	20 – 25°C
3 weeks	at	4 – 8°C
1 year	at	-20°C

Only freeze once. Discard contaminated specimens.

### Calibrators and Controls

DiaSys TruCal U is recommended for calibration. Calibrator values have been made traceable to the NIST Reference Material SRM 682. Use DiaSys TruLab N and P for internal quality control. All target values of the controls are traceable to DiaSys reagent/calibrator system. Quality control must be performed after calibration. Control intervals and limits have to be adapted to the individual requirements of each laboratory. Results must be within the defined ranges. Follow the relevant legal requirements and guidelines. Each laboratory should establish corrective action in case of deviations in control recovery.

	Cat. No.	Kit size
TruCal U	5 9100 99 10 063	20 x 3 mL
	5 9100 99 10 064	6 x 3 mL
TruLab N	5 9000 99 10 062	20 x 5 mL
	5 9000 99 10 061	6 x 5 mL
TruLab P	5 9050 99 10 062	20 x 5 mL
	5 9050 99 10 061	6 x 5 mL

### Performance Characteristics

Measuring range up to 1000 µg/dL. In case of higher concentrations re-measure samples after manual dilution with NaCl solution (9 g/L) or use rerun function.	
Limit of detection**	5 µg/dL
Onboard stability	6 weeks
Calibration stability	6 weeks

Interference by	Interferences ≤ 10% up to	Analyte concentration [µg/dL]
<b>Bilirubin</b> (conjugated)	60 mg/dL	257
<b>Bilirubin</b> (unconjugated)	60 mg/dL	251
<b>Copper</b>	200 µg/dL	97.1
<b>Hemolysis</b>	50 mg/dL	154
<b>Lipemia</b> (triglycerides)	2000 mg/dL	81.1
<b>Zinc</b>	400 µg/dL	95.7

For further information on interfering substances, refer to the literature [5,6].

Precision			
Repeatability (n=20)	Sample 1	Sample 2	Sample 3
Mean [ $\mu\text{g/dL}$ ]	87.2	175	277
CV [%]	1.27	0.741	0.403
Between day (n=20)	Sample 1	Sample 2	Sample 3
Mean [ $\mu\text{g/dL}$ ]	90.5	174	279
CV [%]	0.968	1.20	0.633
Method comparison (n=111)			
Test x	DiaSys Iron FS Ferene (Hitachi 911)		
Test y	DiaSys Iron FS Ferene (respons <sup>®</sup> 920)		
Slope	1.01		
Intercept	3.07 $\mu\text{g/dL}$		
Coefficient of correlation	0.998		

\*\* lowest measurable concentration which can be distinguished from zero; mean + 3 SD (n = 20) of an analyte free specimen.

### Conversion Factor

Iron [ $\mu\text{g/dL}$ ] x 0.1791 = Iron [ $\mu\text{mol/L}$ ]

### Reference Range [7]

	[ $\mu\text{g/dL}$ ]	[ $\mu\text{mol/L}$ ]
<b>Children</b>		
2 weeks	63 – 201	11 – 36
6 months	28 – 135	5 – 24
12 months	35 – 155	6 – 28
2 – 12 years	22 – 135	4 – 24
<b>Women</b>		
25 years	37 – 165	6.6 – 29.5
40 years	23 – 134	4.1 – 24.0
60 years	39 – 149	7.0 – 26.7
<b>Pregnant women</b>		
12th week	42 – 177	7.6 – 31.6
At delivery	25 – 137	4.5 – 24.5
6 weeks post partum	16 – 150	2.9 – 26.9
<b>Men</b>		
25 years	40 – 155	7.2 – 27.7
40 years	35 – 168	6.3 – 30.1
60 years	40 – 120	7.2 – 21.5

Each laboratory should check if the reference ranges are transferable to its own patient population and determine own reference ranges if necessary.

### Literature

1. Wick et al, Clinical Aspects and Lab Iron Metabolism, Anemias. Novel concepts, Springer, 5th ed. Wien New York 2003 p141 – 147.
2. Fairbanks VF, Klee GG. Biochemical aspects of hematology. In: Burtis CA, Ashwood ER, editors. Tietz Textbook of Clinical Chemistry. 3rd ed. Philadelphia: W.B Saunders Company; 1999. p. 1642-1710.
3. Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: mechanisms, detection and prevention. ClinChemLabMed 2007;45(9):1240-1243.
4. Guder WG, Zawta B et al. The Quality of Diagnostic Samples. 1st ed. Darmstadt: GIT Verlag; 2001; p. 34-5.
5. Young DS. Effects of Drugs on Clinical Laboratory Tests. 5th ed. Volume 1 and 2. Washington, DC: The American Association for Clinical Chemistry Press 2000.
6. Young DS. Effects on Clinical Laboratory Tests - Drugs Disease, Herbs & Natural Products, <https://clinf.wiley.com/aaccweb/aacc/>, accessed in March 2024. Published by AACC Press and John Wiley and Sons, Inc.
7. Thomas L. Clinical Laboratory Diagnostics [Internet]. Prof. Lothar Thomas; 2024 [cited 2024 May 13]. <https://www.clinical-laboratory-diagnostics.com>

Additions and/or changes in the document are highlighted in grey. Deletions are communicated via customer info by stating the edition no. of the package insert/instruction for use.



DiaSys Diagnostic Systems GmbH  
Alte Strasse 9 65558 Holzheim  
Germany  
[www.diasys-diagnostics.com](http://www.diasys-diagnostics.com)

\* Fluid Stable

## Iron FS Ferene

### Application for serum and plasma

Test Details		Test Volumes		Reference Ranges	
Test	: FE			Auto Rerun	<input type="checkbox"/>
Report Name	: IRON			Online Calibration	<input type="checkbox"/>
Unit	: µg/dL	Decimal Places	: 2	Cuvette Wash	<input type="checkbox"/>
Wavelength-Primary	: 578	Secondary	: 700	Total Reagents	: 2
Assay Type	: 2-Point	Curve Type	: Linear	Reagent R1	: FE R1
M1 Start	: 15	M1 End	: 15	Reagent R2	: FE R2
M2 Start	: 33	M2 End	: 33	<b>Consumables/Calibrators:</b>	
Sample Replicates	: 1	Standard Replicates	: 3	Blank/Level 0	: 0
Control Replicates	: 1	Control Interval	: 0	Calibrator 1	: *
Reaction Direction	: Increasing	React. Abs. Limit	: 0.00		
Prozone Limit %	: 0	Prozone Check	: Lower		
Linearity Limit %	: 0	Delta Abs./Min.	: 0.00		
Technical Minimum	: 5.00	Technical Maximum	: 1000.00		
Y = aX + b	a = 1.00	b =	: 0.00		

\* Enter calibrator value

Test Details		Test Volumes		Reference Ranges	
Test	: FE				
Sample Type	: Serum				
<b>Sample Volumes</b>				<b>Sample Types</b>	
Normal	: 11.00 µL	Dilution Ratio	: 1 X	<input checked="" type="checkbox"/> Serum <input type="checkbox"/> Urine <input type="checkbox"/> CSF <input checked="" type="checkbox"/> Plasma <input type="checkbox"/> Whole Blood <input type="checkbox"/> Other	
Increase	: 15.00 µL	Dilution Ratio	: 1 X		
Decrease	: 8.00 µL	Dilution Ratio	: 1 X		
Standard Volume	: 11.00 µL				
<b>Reagent Volumes and Stirrer Speed</b>					
RGT-1 Volume	: 180 µL	R1 Stirrer Speed	: Medium		
RGT-2 Volume	: 45 µL	R2 Stirrer Speed	: High		

Test Details		Test Volumes		Reference Ranges	
Test	: FE				
Sample Type	: Serum				
Reference Range	: DEFAULT				
Category	: Male				
<b>Reference Range</b>				<b>Sample Types</b>	
		Lower Limit		Upper Limit	
		(µg/dL)		(µg/dL)	
Normal	: 35.00			168.00	<input checked="" type="checkbox"/> Serum <input type="checkbox"/> Urine <input type="checkbox"/> CSF <input checked="" type="checkbox"/> Plasma <input type="checkbox"/> Whole Blood <input type="checkbox"/> Other
Panic	: 0.00			0.00	