

Lactate FS*

Order Information

Cat. No.

1 4001 99 10 921

Kit size



480 (4 x 120)

Intended Use


Diagnostic reagent for quantitative in vitro determination of lactate in human heparin plasma on automated respons[®]940.

Summary

Lactate is the final product of the anaerobic glycolysis and serves as indicator for the oxygen status in cellular tissues. Increased lactate levels in blood occur in anoxia due to shock, congestive heart failure, intoxication and thiamine deficiency [1]. Therefore, lactate is measured in intensive care medicine. Lactate measurement is recommended in patients with sepsis. An elevated lactate level is part of the Sepsis 3 definition of septic shock [2]. As metabolic variable for the capability of the muscles, lactate determination is used in evaluation of the training status in athletes [3].

Method

Enzymatic UV test with lactate dehydrogenase (LDH)



In the presence of NAD, lactate is converted by the lactate dehydrogenase. This procedure releases NADH, which is measured at 340 nm. The absorbance of the produced NADH is proportional to the lactate concentration in the sample.

Reagents

Components and Concentrations

R1: Buffer	pH 9.0	500 mmol/L
LDH		≥ 25 kU/L
R2: NAD		20 mmol/L

Storage and Stability

Reagents are stable up to the date of expiry indicated on the kit, if stored at 2 – 8°C and contamination is avoided. Do not freeze and protect from light.

The open-vial stability of the reagent is 15 months until expiry date.

Warnings and Precautions

- Components contained in Lactate FS are classified according to EC regulation 1272/2008 (CLP) as follows:



⚠ Reagent 1: Danger. Contains 1,3-Diaminopropan-2-ol. H315 Causes skin irritation. H318 Causes serious eye damage. P264 Wash hands and face thoroughly after handling. P280 Wear protective gloves/protective clothing/eye protection. P305+P351+P338 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. P310 Immediately call a POISON CENTER/doctor.

- Reagent 1 contains sodium azide (0.95 g/L) as preservative. Do not swallow! Avoid contact with skin and mucous membranes.
- Reagent 1 contains material of biological origin. Handle the product as potentially infectious according to universal precautions and good clinical laboratory practice.
- In very rare cases, samples of patients with gammopathy might give falsified results [4].
- In case of product malfunction or altered appearance that could affect the performance, contact the manufacturer.
- Any serious incident related to the product must be reported to the manufacturer and the competent authority of the Member State where the user and/or patient is located.
- Please refer to the safety data sheets (SDS) and take the necessary precautions for the use of laboratory reagents. For diagnostic purposes, the results should always be assessed with the patient's medical history, clinical examinations and other findings.
- For professional use only.

Waste Management

Refer to local legal requirements for chemical disposal regulations as stated in the relevant SDS to determine the safe disposal.

Warning: Handle waste as potentially biohazardous material. Dispose of waste according to accepted laboratory instructions and procedures.

Reagent Preparation

The reagents are ready to use. The bottles are placed directly into the reagent rotor.

Materials Required

General laboratory equipment

Specimen

Human heparin plasma

Only use suitable tubes or collection containers for specimen collection and preparation.

When using primary tubes, follow the manufacturer's instructions.

Use glycolytic inhibitors e.g. fluoride/oxalate or fluoride/heparin as anticoagulants.

Stability [5]:

6 days	at	20 – 25°C
14 days	at	2 – 8°C
1 month	at	-20°C

Only freeze once. Discard contaminated specimens.

Calibrators and Controls

DiaSys TruCal U is recommended for calibration. Calibrator values have been made traceable to a primary standard. Use DiaSys TruLab N and P for internal quality control. Quality control must be performed after calibration. Control intervals and limits have to be adapted to the individual requirements of each laboratory. Results must be within the defined ranges. Follow the relevant legal requirements and guidelines. Each laboratory should establish corrective action in case of deviations in control recovery.

	Cat. No.	Kit size
TruCal U	5 9100 99 10 063	20 x 3 mL
	5 9100 99 10 064	6 x 3 mL
TruLab N	5 9000 99 10 062	20 x 5 mL
	5 9000 99 10 061	6 x 5 mL
TruLab P	5 9050 99 10 062	20 x 5 mL
	5 9050 99 10 061	6 x 5 mL

Performance Characteristics

Measuring range from 1.1 mg/dL up to 150 mg/dL. Linearity < 1.6 mg/dL is given with ± 0.6 mg/dL, between 1.6 mg/dL to 10 mg/dL within ± 10%, at > 10 mg/dL within ± 5%. In case of higher concentrations re-measure samples after manual dilution with NaCl solution (9 g/L) or use rerun function.	
Limit of detection**	1.1 mg/dL
Limit of quantitation**	1.1 mg/dL
Onboard stability (with chimney)	6 weeks
Calibration stability (with chimney)	3 weeks

Interference by	Interferences ≤ 10% up to	Analyte concentration [mg/dL]
Ascorbic acid	30 mg/dL	8.99
	30 mg/dL	13.8
Bilirubin (conjugated)	75 mg/dL	9.60
	75 mg/dL	15.6
Bilirubin (unconjugated)	48 mg/dL	8.90
	65 mg/dL	14.8
Dopamine	12 mg/L	9.47
	12 mg/L	15.2
Glycolic acid	1200 mg/L	8.88
	1200 mg/L	14.2
Hemolysis	1000 mg/dL	9.01
	1300 mg/dL	14.8
Levodopa	25 mg/L	7.57
	25 mg/L	12.7
Lipemia (triglycerides)	1500 mg/dL	9.36
	1800 mg/dL	14.9
Methyldopamine	12 mg/L	9.04
	12 mg/L	14.3

For further information on interfering substances, refer to the literature [6,7].

Precision			
Repeatability (n=20)	Sample 1	Sample 2	Sample 3
Mean [mg/dL]	4.59	20.9	62.5
CV [%]	2.94	1.34	1.75
Within-laboratory (n=80)	Sample 1	Sample 2	Sample 3
Mean [mg/dL]	4.47	22.8	72.6
CV [%]	4.27	1.53	1.94

Method comparison (n=131)	
Test x	Competitor Lactate (cobas c 501)
Test y	DiaSys Lactate FS (respons [®] 940)
Slope	0.853
Intercept	0.333 mg/dL
Coefficient of correlation	0.997

** according to CLSI document EP17-A2, Vol. 32, No. 8

Conversion Factor

Lactate [mg/dL] x 0.1109 = Lactate [mmol/L]

Reference Range [8]

Plasma

Venous 4.5 – 19.8 mg/dL 0.5 – 2.2 mmol/L
Arterial 4.5 – 14.4 mg/dL 0.5 – 1.6 mmol/L

Each laboratory should check if the reference ranges are transferable to its own patient population and determine own reference ranges if necessary.

Literature

- David B. Sacks, M.B., Ch.B., F.R.C.Path. Carbohydrates In: Burtis CA, Ashwood ER, editors. Tietz Textbook of Clinical Chemistry. 3rd ed. Philadelphia: W.B Saunders Company; 1999. p. 787–789.
- Singer M, Deutschman CS, Seymour CW et al. The Third International Consensus Definitions for Sepsis and Septic Shock (Sepsis-3). JAMA 2016; 315,8: 801-10
- Thomas L. Clinical Laboratory Diagnostics [Internet]. Prof. Lothar Thomas; 2020 [cited 2023 Nov 21]. Available from: <https://www.clinical-laboratory-diagnostics.com/>
- Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: Mechanism, detection and prevention. Clin Chem Lab Med 2007; 45(9): 1240-1243.
- Guder WG, Zawta B et al. Quality of Diagnostic Samples. German Society of Clinical Chemistry. 3rd ed.2010; p. 52-53.

- Young DS. Effects of Drugs on Clinical Laboratory Tests. 5th ed. Volume 1 and 2. Washington, DC: The American Association for Clinical Chemistry Press 2000.
- Young DS. Effects on Clinical Laboratory Tests - Drugs Disease, Herbs & Natural Products, <https://clinfx.wiley.com/aaccweb/aacc/>, accessed in November 2023. Published by AACC Press and John Wiley and Sons, Inc.
- Section I – General Clinical Tests In: Tietz NW, editor. Clinical Guide to Laboratory Tests. 3rd ed. Philadelphia: Saunders; 1995. p. 382-3.

Additions and/or changes in the document are highlighted in grey. Deletions are communicated via customer info by stating the edition no. of the package insert/instruction for use.



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim
Germany
www.diasys-diagnostics.com

* Fluid Stable

Lactate FS

Application for serum and plasma

Test Details		Test Volumes		Reference Ranges	
Test	: LACT			Auto Rerun	<input type="checkbox"/>
Report Name	: Lactate			Online Calibration	<input type="checkbox"/>
Unit	: mg/dL	Decimal Places	: 2	Cuvette Wash	<input type="checkbox"/>
Wavelength-Primary	: 340	Secondary	: 700	Special Diluent	<input type="checkbox"/>
Assay Type	: 2-Point	Curve Type	: Linear	Warn after	: 20
M1 Start	: 24	M1 End	: 24	Reagents Used	: 2
M2 Start	: 57	M2 End	: 57	Reagent R1	LACT R1
Sample Replicates	: 1	Standard Replicates	: 2	Reagent R2	LACT R2
Control Replicates	: 1	Control Interval	: 0	Consumables/Calibrators:	
Reaction Direction	: Increasing	React. Abs. Limit	: 0.0000	Blank /Level 0	0
Prozone Limit %	: 0	Prozone Check	: Lower	Calibrator 1	*
Linearity Limit %	: 0	Delta Abs./Min.	: 0.0000	Calibrator 2	
Technical Minimum	: 1.1000	Technical Maximum	: 150.0000	Calibrator 3	
Y = aX + b a=	: 1.0000	b=	: 0.0000	Calibrator 4	
Reagent Abs Min	: 0.0000	Reagent Abs Max	: 0.0000	Calibrator 5	

Test Details		Test Volumes		Reference Ranges	
Test	: LACT				
Sample Type	: Serum				
Sample Volumes				Sample Types	
Normal	: 3.00 μ L	Dilution Ratio	: 1 X	<input checked="" type="checkbox"/> Serum <input type="checkbox"/> Urine <input type="checkbox"/> CSF <input checked="" type="checkbox"/> Plasma <input type="checkbox"/> Whole Blood <input type="checkbox"/> Other	
Increase	: 6.00 μ L	Dilution Ratio	: 1 X		
Decrease	: 2.00 μ L	Dilution Ratio	: 1 X		
Standard Volume	: 3.00 μ L				
Reagent Volumes and Stirrer Speed					
RGT-1 Volume	: 180.00 μ L	R1 Stirrer Speed	: Medium		
RGT-2 Volume	: 45.00 μ L	R2 Stirrer Speed	: High		

Test Details		Test Volumes		Reference Ranges	
Test	: LACT				
Sample Type	: Serum				
Reference Range	: DEFAULT				
Category	: Male				
Reference Range				Sample Types	
	Lower Limit (mg/dL)	Upper Limit (mg/dL)			
Normal	: #	: #	<input checked="" type="checkbox"/> Serum <input type="checkbox"/> Urine <input type="checkbox"/> CSF <input checked="" type="checkbox"/> Plasma <input type="checkbox"/> Whole Blood <input type="checkbox"/> Other		
Panic	: #	: #			

* Enter calibrator value
Editable by user