


Antistreptolysin O FS*

Order Information

Cat. No.	Kit size
1 7012 99 10 921	 400 (4 x 100)

Intended Use

Diagnostic reagent for quantitative in vitro determination of antistreptolysin O (ASO) in human serum on automated respons[®]940.

Summary

Antistreptolysins (ASL) are specific antibodies to extracellular products of *Streptococcus pyogenes* (Group A streptococcus: GAS), among which antistreptolysin O (ASO) is the one most used for clinical laboratory evaluation [1]. Antistreptolysin O determination provides useful information for diagnosis and monitoring of human streptococcal infections such as in tonsillitis, otitis, erysipela, scarlet fever as well as connected diseases like rheumatic fever or glomerulonephritis. Antibodies against streptolysin O can be detected 1 – 3 weeks after infection with maximum levels reached at 3 – 6 weeks [2]. Pathological ASO values always indicate the presence of a streptococcal infection whereas a negative result cannot exclude an existing or preceding GAS infection [3].

Method

Particle enhanced immunoturbidimetric test

Determination of ASO concentration by photometric measurement of the antigen antibody reaction of latex particles coated with streptolysin O and antibodies to streptolysin O present in the sample.

Reagents

Components and Concentrations

R1:	Phosphate buffer	pH 7.0	100 mmol/L
	NaCl		150 mmol/L
R2:	Glycine buffer	pH 8.0	100 mmol/L
	NaCl		150 mmol/L
	Latex particles coated with streptolysin O		< 1%

Storage and Stability

Reagents are stable up to the date of expiry indicated on the kit, if stored at 2 – 8°C and contamination is avoided. Do not freeze.

The open-vial stability of the reagent is 18 months until expiry date.

Warnings and Precautions

- The reagents contain sodium azide (0.95 g/L) as preservative. Do not swallow! Avoid contact with skin and mucous membranes.
- The reagents contain material of biological origin. Handle the product as potentially infectious according to universal precautions and good clinical laboratory practice.
- In very rare cases, samples of patients with gammopathy might give falsified results [4].
- In case of product malfunction or altered appearance that could affect the performance, contact the manufacturer.
- Any serious incident related to the product must be reported to the manufacturer and the competent authority of the Member State where the user and/or patient is located.
- Please refer to the safety data sheets (SDS) and take the necessary precautions for the use of laboratory reagents. For diagnostic purposes, the results should always be assessed with the patient's medical history, clinical examinations and other findings.
- For professional use only.

Waste Management

Refer to local legal requirements for chemical disposal regulations as stated in the relevant SDS to determine the safe disposal.

Warning: Handle waste as potentially biohazardous material. Dispose of waste according to accepted laboratory instructions and procedures.

Reagent Preparation

The latex enhanced reagent 2 in respons[®] bottle has to be mixed up by successive inversions before first use. Avoid formation of foam.

Materials Required

General laboratory equipment

Specimen

Human serum

Only use suitable tubes or collection containers for specimen collection and preparation.

When using primary tubes, follow the manufacturer's instructions.

Stability [5]:

2 days	at	20 – 25°C
8 days	at	4 – 8°C
6 months	at	-20°C

Only freeze once. Discard contaminated specimens.

Calibrators and Controls

DiaSys TruCal ASO is recommended for calibration. Calibrator values have been made traceable to a commercially available standard material, traceable to the first International standard as ASL reference standard. Use TruLab Protein Level 1 and Level 2 for internal quality control. All target values of the controls are traceable to DiaSys reagent/calibrator system. Quality control must be performed after calibration. Control intervals and limits have to be adapted to the individual requirements of each laboratory. Results must be within the defined ranges. Follow the relevant legal requirements and guidelines. Each laboratory should establish corrective action in case of deviations in control recovery.

	Cat. No.	Kit size
TruCal ASO	1 7010 99 10 059	5 x 1 mL
TruLab Protein Level 1	5 9500 99 10 046	3 x 1 mL
TruLab Protein Level 2	5 9510 99 10 046	3 x 1 mL

Performance Characteristics

Measuring range from 20 IU/mL up to 750 IU/mL, depending on the concentration of the highest calibrator. Linearity < 50 IU/mL is given with ± 12 IU/mL, between 50 IU/mL to 100 IU/mL within ± 10%, at > 100 IU/mL within ± 5%. In case of higher concentrations re-measure samples after manual dilution with NaCl solution (9 g/L) or use rerun function.	
Limit of detection**	20 IU/mL
Limit of quantitation**	20 IU/mL
No prozone effect up to 3000 IU/mL.	
Onboard stability	12 weeks
Calibration stability	2 weeks

Interference by	Interferences ≤ 10% up to	Analyte concentration [IU/mL]
Acetaminophen	18 mg/dL	97.3
	18 mg/dL	332
Acetylsalicylic acid	3.6 mg/dL	112
	3.6 mg/dL	312
Bilirubin (conjugated)	70 mg/dL	102
	70 mg/dL	298
Bilirubin (unconjugated)	70 mg/dL	96.7
	70 mg/dL	237
Hemolysis	600 mg/dL	97.2
	600 mg/dL	313
Ibuprofen	25 mg/dL	105
	25 mg/dL	323
Lipemia (triglycerides)	2400 mg/dL	101
	2400 mg/dL	303
Rheumatoid factor	400 IU/mL	87.2
	720 IU/mL	290

For further information on interfering substances, refer to the literature [6,7].

Precision			
Repeatability (n=20)	Sample 1	Sample 2	Sample 3
Mean [IU/mL]	134	204	512
CV [%]	0.753	1.29	1.88
Within-laboratory (n=80)	Sample 1	Sample 2	Sample 3
Mean [IU/mL]	94.0	199	675
CV [%]	2.71	2.47	3.72

Method comparison (n=97)	
Test x	Competitor Antistreptolysin O (cobas c 501)
Test y	DiaSys Antistreptolysin O FS (respons [®] 940)
Slope	0.956
Intercept	4.02 IU/mL
Coefficient of correlation	0.965

** according to CLSI document EP17-A2, Vol. 32, No. 8

Reference Range [8]

Adults	≤ 200 IU/mL
Children < 6 years	≤ 150 IU/mL
Children 6 – 18 years	≤ 200 – 240 IU/mL

Each laboratory should check if the reference ranges are transferable to its own patient population and determine own reference ranges if necessary.

Literature

1. Bisno AL. Group A infections and acute rheumatic fever. N Engl J Med 1991;325:783-93.
2. Curtis GD, Kraak WA, Mitchell RG. Comparison of latex and haemolysin tests for determination of anti-streptolysin O (ASO) antibodies. J Clin Pathol 1988; 41:1 331-3.
3. Stevens DL. Invasive Group A streptococcus infections. Clin Infect Dis 1992;14:2-11.
4. Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: mechanisms, detection and prevention. ClinChemLabMed 2007;45(9):1240-1243.
5. Guder WG, da Fonseca-Wollheim F, Heil W, et al. The Quality of Diagnostic Samples, German United Society for Clinical Chemistry and Laboratory Medicine. 3rd ed; 2010.
6. Young DS. Effects of Drugs on Clinical Laboratory Tests. 5th ed. Volume 1 and 2. Washington, DC: The American Association for Clinical Chemistry Press 2000.
7. Young DS. Effects on Clinical Laboratory Tests - Drugs Disease, Herbs & Natural Products, <https://clinfo.wiley.com/>

aaccweb/aacc/, accessed in March 2024. Published by AACC Press and John Wiley and Sons, Inc.

8. Thomas L. Clinical Laboratory Diagnostics [Internet]. Prof. Lothar Thomas; 2024 [cited 2024 May 22]. <https://www.clinical-laboratory-diagnostics.com>

Additions and/or changes in the document are highlighted in grey. Deletions are communicated via customer info by stating the edition no. of the package insert/instruction for use.



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim
Germany
www.diasys-diagnostics.com

* Fluid Stable

Antistreptolysin O FS

Application for serum

Test Details		Test Volumes		Reference Ranges	
Test	: ASO			Auto Rerun	<input type="checkbox"/>
Report Name	: Antistreptolysin O			Online Calibration	<input type="checkbox"/>
Unit	: IU/mL	Decimal Places	: 1	Cuvette Wash	<input type="checkbox"/>
Wavelength-Primary	: 570	Secondary	: 0	Special Diluent	<input type="checkbox"/>
Assay Type	: 2-Point	Curve Type	: Cubic Spline	Warn after	: 20
M1 Start	: 29	M1 End	: 29	Reagents Used	: 2
M2 Start	: 57	M2 End	: 57	Reagent R1	ASO R1
Sample Replicates	: 1	Standard Replicates	: 2	Reagent R2	ASO R2
Control Replicates	: 1	Control Interval	: 0	Consumables/Calibrators:	
Reaction Direction	: Increasing	React. Abs. Limit	: 0.0000	Blank /Level 0	0
Prozone Limit %	: 97	Prozone Check	: Lower	Calibrator 1	*
Linearity Limit %	: 0	Delta Abs./Min.	: 0.0000	Calibrator 2	*
Technical Minimum	: **	Technical Maximum	: **	Calibrator 3	*
Y = aX + b a=	: 1.0000	b=	: 0.0000	Calibrator 4	*
Reagent Abs Min	: 0.0000	Reagent Abs Max	: 0.0000	Calibrator 5	*

Test Details		Test Volumes		Reference Ranges	
Test	: ASO				
Sample Type	: Serum				
Sample Volumes				Sample Types	
Normal	: 2.00 μ L	Dilution Ratio	: 1 X	<input checked="" type="checkbox"/> Serum <input type="checkbox"/> Urine <input type="checkbox"/> CSF <input type="checkbox"/> Plasma <input type="checkbox"/> Whole Blood <input type="checkbox"/> Other	
Increase	: 6.00 μ L	Dilution Ratio	: 1 X		
Decrease	: 2.00 μ L	Dilution Ratio	: 2 X		
Standard Volume	: 2.00 μ L				
Reagent Volumes and Stirrer Speed					
RGT-1 Volume	: 160.00 μ L	R1 Stirrer Speed	: High		
RGT-2 Volume	: 32.00 μ L	R2 Stirrer Speed	: High		

Test Details		Test Volumes		Reference Ranges	
Test	: ASO				
Sample Type	: Serum				
Reference Range	: DEFAULT				
Category	: Male				
Reference Range				Sample Types	
	Lower Limit (IU/mL)	Upper Limit (IU/mL)			
Normal	: #	: #	<input checked="" type="checkbox"/> Serum <input type="checkbox"/> Urine <input type="checkbox"/> CSF <input type="checkbox"/> Plasma <input type="checkbox"/> Whole Blood <input type="checkbox"/> Other		
Panic	: #	: #			

* Enter calibrator value

** Technical limits are automatically defined by the software via the upper and lower calibrator level

Editable by user