

# Alkaline phosphatase FS\*


IFCC mod. 37°C

## Order Information

### Cat. No.

1 0441 99 10 920

### Kit size

 800 (4 x 200)

## Intended Use


Diagnostic reagent for quantitative in vitro determination of alkaline phosphatase (AP) in serum or plasma on respons<sup>®</sup>910.

## Summary

Alkaline phosphatase (AP), a hydrolytic enzyme acting optimally at alkaline pH, exists in blood in numerous distinct forms which originate mainly from bone and liver, but also from other tissues as kidney, placenta, testes, thymus, lung and tumors. Physiological increases are found during bone growth in childhood and in pregnancy, while pathological increases are largely associated with hepatobiliary and bone diseases. In hepatobiliary disease they indicate obstruction of the bile ducts as in cholestasis caused by gall stones, tumors or inflammation. Elevated activities are also observed in infectious hepatitis. In bone diseases elevated AP activities originate from increased osteoblastic activity as in Paget's disease, osteomalacia (rickets), bone metastases and hyperparathyroidism. [1,2]

## Method

Kinetic photometric test, according to the International Federation of Clinical Chemistry and Laboratory Medicine (IFCC) [modif.] [3].



## Reagents

### Components and Concentrations

<b>R1:</b>	2-Amino-2-methyl-1-propanol	pH 10.4	1.1 mol/L
	Magnesium acetate		2 mmol/L
	Zinc sulphate		0.5 mmol/L
	HEDTA		2.5 mmol/L
<b>R2:</b>	p-Nitrophenylphosphate		80 mmol/L

## Storage and Stability

The reagents are stable up to the end of the indicated month of expiry, if stored at 2 – 8°C and contamination is avoided. Do not freeze the reagents and protect them from light.

DiaSys respons containers provide protection from light.

## Warnings and Precautions

- The reagents contain sodium azide (0.95 g/L) as preservative. Do not swallow! Avoid contact with skin and mucous membranes.
- During the reaction, p-nitrophenol is produced which is poisonous when inhaled, swallowed or absorbed through skin. If the reaction mixture comes in contact with skin or mucous membranes wash copiously with water!
- In very rare cases, samples of patients with gammopathy might give falsified results [4].
- To avoid carryover interference, please take care of efficient washing especially after use of interfering reagents. Please refer to the DiaSys respons<sup>®</sup>910 Carryover Pair Table. Carryover pairs and automated washing steps with the recommended cleaning solution can be specified in the system software. Please refer to the user manual.
- Please refer to the safety data sheets and take the necessary precautions for the use of laboratory reagents. For diagnostic purposes, the results should always be assessed with the patient's medical history, clinical examinations and other findings.
- For professional use only.

## Waste Management

Refer to local legal requirements.

## Reagent Preparation

The reagent is ready to use. The bottles are placed directly into the reagent rotor.

## Materials Required

General laboratory equipment

## Specimen

Serum or heparin plasma

Do not use hemolytic samples.

### Stability [5]:

7 days	at	20 – 25°C
7 days	at	4 – 8°C
2 months	at	-20°C

Only freeze once. Discard contaminated specimens.

## Calibrators and Controls

DiaSys TruCal U is recommended for calibration. This method is traceable to the molar extinction coefficient. Use DiaSys TruLab N and P for internal quality control. Each laboratory should establish corrective action in case of deviations in control recovery.

	Cat. No.	Kit size		
TruCal U	5 9100 99 10 063	20	x	3 mL
	5 9100 99 10 064	6	x	3 mL
TruLab N	5 9000 99 10 062	20	x	5 mL
	5 9000 99 10 061	6	x	5 mL
TruLab P	5 9050 99 10 062	20	x	5 mL
	5 9050 99 10 061	6	x	5 mL

## Performance Characteristics

Exemplary data mentioned below may slightly differ in case of deviating measurement conditions.

Measuring range up to 1400 U/L. In case of higher concentrations re-measure samples after manual dilution with NaCl solution (9 g/L) or use rerun function.	
Limit of detection**	3 U/L
Onboard stability	7 days
Calibration stability	7 days

Interfering substance	Interferences ≤ 10% up to	Analyte concentration [U/L]
<b>Ascorbate</b>	30 mg/dL	154
<b>Hemoglobin</b>	100 mg/dL	74.2
	100 mg/dL	310
<b>Bilirubin (conjugated)</b>	80 mg/dL	95.2
	80 mg/dL	182
<b>Bilirubin (unconjugated)</b>	70 mg/dL	94.9
	70 mg/dL	188
<b>Lipemia (triglycerides)</b>	2200 mg/dL	98.6
	2200 mg/dL	202

For further information on interfering substances refer to Young DS [6].

Precision			
Within run (n=20)	Sample 1	Sample 2	Sample 3
Mean [U/L]	76.6	122	229
CV [%]	1.62	1.44	1.81
Between day (n=20)	Sample 1	Sample 2	Sample 3
Mean [U/L]	73.7	127	213
CV [%]	4.04	4.83	3.13

Method comparison (n=117)	
Test x	DiaSys Alkaline phosphatase FS (Hitachi 911)
Test y	DiaSys Alkaline phosphatase FS (respons <sup>®</sup> 910)
Slope	1.049
Intercept	-4.67 U/L
Coefficient of correlation	0.9996

\*\* according to CLSI document EP17-A, Vol. 24, No. 34

## Conversion Factor

AP [U/L] x 0.0167 = AP [µkat/L]

## Reference Range

Adults [7]		
Women	35 – 104 [U/L]	0.58 – 1.74 µkat/L
Men	40 – 129 [U/L]	0.67 – 2.15 µkat/L

Adults [8]		
Women	35 – 105 [U/L]	0.58 – 1.75 µkat/L
Men	40 – 130 [U/L]	0.67 – 2.17 µkat/L

Children [9]				
	Female [U/L]	Male [U/L]	Female [µkat/L]	Male [µkat/L]
1 – 30 day(s)	48 – 406	75 – 316	0.80 – 6.77	1.25 – 5.27
1 month – 1 year	124 – 341	82 – 383	2.07 – 5.68	1.37 – 6.38
1 – 3 year(s)	108 – 317	104 – 345	1.80 – 5.28	1.73 – 5.75
4 – 6 years	96 – 297	93 – 309	1.60 – 4.95	1.55 – 5.15
7 – 9 years	69 – 325	86 – 315	1.15 – 5.42	1.43 – 5.25
10 – 12 years	51 – 332	42 – 362	0.85 – 5.53	0.70 – 6.03
13 – 15 years	50 – 162	74 – 390	0.83 – 2.70	1.23 – 6.50
16 – 18 years	47 – 119	52 – 171	0.78 – 1.98	0.87 – 2.85

Each laboratory should check if the reference ranges are transferable to its own patient population and determine own reference ranges if necessary.

## Literature

1. Thomas L. Clinical Laboratory Diagnostics. 1st ed. Frankfurt: TH-Books Verlagsgesellschaft; 1998. p. 36-46.
2. Moss DW, Henderson AR. Clinical enzymology. In: Burtis CA, Ashwood ER, editors. Tietz Textbook of Clinical Chemistry. 3rd ed. Philadelphia: W.B Saunders Company; 1999. p. 617-721.
3. IFCC primary reference procedures for the measurement of catalytic activity concentrations of enzymes at 37°C. Part 9: Reference procedure for the measurement of catalytic concentration of alkaline phosphatase; Clin Chem Lab Med 2011;49(9).
4. Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: mechanisms, detection and prevention. ClinChemLabMed 2007;45(9):1240-1243.
5. Guder WG, Zawta B et al. The Quality of Diagnostic Samples. 1st ed. Darmstadt: GIT Verlag; 2001; p. 14-5.
6. Young DS. Effects of Drugs on Clinical Laboratory Tests. 5th ed. Volume 1 and 2. Washington, DC: The American Association for Clinical Chemistry Press 2000.
7. Abicht K et al. Multicenter evaluation of new GGT and ALP reagents with new reference standardization and determination of 37 °C reference intervals. Clin Chem Lab Med 2001; 39 (Suppl.): S 346 [abstract].
8. Thomas L, Müller M, Schumann G, Weidemann G et al. Consensus of DGKL and VDGH for interim reference intervals on enzymes in serum. J Lab Med 2005;29:301-308.
9. Soldin JS, Brugnara C., Wong CE. In: MJ Hicks, editor. Pediatric reference intervals. 6th ed. Washington: AACC Press, 2007. p. 11.

## Alkaline phosphatase FS IFCC 37 °C

### Application for serum and plasma samples

This application was set up and evaluated by DiaSys. It is based on the standard equipment at that time and does not apply to any equipment modifications undertaken by unqualified personnel.

Identification	
This method is usable for analysis:	Yes
Twin reaction:	No
Name:	AP
Shortcut:	
Reagent barcode reference:	014
Host reference:	

Technic	
Type:	Linear kinetic
First reagent:[ $\mu$ L]	160
Blank reagent	Yes
Sensitive to light	
Second reagent:[ $\mu$ L]	40
Blank reagent	Yes
Sensitive to light	
Main wavelength:[nm]	405
Secondary wavelength:[nm]	700
Polychromatic factor:	1.000
1 st reading time [min:sec]	6:48
Last reading time [min:sec]	10:00
Reaction way:	Increasing
Linear Kinetics	
Substrate depletion: Absorbance li	1.4000
Linearity: Maximum deviation [%]	100
Fixed Time Kinetics	
Substrate depletion: Absorbance limit	
Endpoint	
Stability: Largest remaining slope	
Prozone Limit [%]	

Reagents	
Decimals	
Units	

Sample	
Diluent	DIL A (NaCl)
Hemolysis:	
Agent [ $\mu$ L]	0 (no hemolysis)
Cleaner	
Sample [ $\mu$ L]	0
Technical limits	
Concentration technical limits-Lower	3
Concentration technical limits-Upper	1400
SERUM	
Normal volume [ $\mu$ L]	3
Normal dilution (factor)	1
Below normal volume [ $\mu$ L]	6
Below normal dilution (factor)	1
Above normal volume [ $\mu$ L]	3
Above normal dilution (factor)	6
URIN	
Normal volume [ $\mu$ L]	3
Normal dilution (factor)	1
Below normal volume [ $\mu$ L]	6
Below normal dilution (factor)	1
Above normal volume [ $\mu$ L]	3
Above normal dilution (factor)	6
PLASMA	
Normal volume [ $\mu$ L]	3
Normal dilution (factor)	1
Below normal volume [ $\mu$ L]	6
Below normal dilution (factor)	1
Above normal volume [ $\mu$ L]	3
Above normal dilution (factor)	6
CSF	
Normal volume [ $\mu$ L]	3
Normal dilution (factor)	1
Below normal volume [ $\mu$ L]	6
Below normal dilution (factor)	1
Above normal volume [ $\mu$ L]	3
Above normal dilution (factor)	6
Whole blood	
Normal volume [ $\mu$ L]	3
Normal dilution (factor)	1
Below normal volume [ $\mu$ L]	6
Below normal dilution (factor)	1
Above normal volume [ $\mu$ L]	3
Above normal dilution (factor)	6

Results	
Decimals	1
Units	U/L
Correlation factor-Offset	0.000
Correlation factor-Slope	1.000

Range	
Gender	Male
Age	
SERUM	>=40 <=130
URINE	
PLASMA	>=40 <=130
CSF	
Whole blood	
Gender	Female
Age	
SERUM	>=35 <=105
URINE	
PLASMA	>=35 <=105
CSF	
Whole blood	

Contaminants	
Please refer to r910 Carryover Pair Table	

Calibrators details	
Calibrator list	Concentration
Cal. 1/Blank	0
Cal. 2	*
Cal. 3	
Cal. 4	
Cal. 5	
Cal. 6	
	Max delta abs.
Cal. 1	0.002
Cal. 2	0.005
Cal. 3	
Cal. 4	
Cal. 5	
Cal. 6	
Drift limit [%]	0.8

Calculations	
Model	X
Degree	1

\* Enter calibrator value