

SARS-CoV-2 UTAB Sample Dilution Matrix

Order Information

Cat. No. Kit size
1 7501 99 10 021 6 x 25 mL

Method

Sample dilution matrix for determination of total antibodies in COVID-19 with the DiaSys SARS-CoV-2 UTAB FS reagent. SARS-CoV-2 UTAB Sample Dilution Matrix is used for dilution of samples exceeding the measuring range.

Reagents

Components and Concentrations

Phosphate buffer pH 7.4 8.55 mmol/L

Storage

The sample dilution matrix both opened and unopened must be stored at 2 – 8°C. Avoid contamination and protect from light. Do not freeze.

Stability

Unopened: up to the date of expiry indicated on the kit
Opened: 2 months
Proper storage and handling of this product must be observed.

Warnings and Precautions

1. Contains sodium azide (0.95 g/L) as preservative. Do not swallow! Avoid contact with skin and mucous membranes.
2. Contains animal material. Handle the product as potentially infectious according to universal precautions and good clinical laboratory practices.
3. Please refer to the safety data sheets and take the necessary precautions for the use of laboratory reagents. For diagnostic purposes, the results should always be assessed with the patient's medical history, clinical examinations and other findings.
4. For professional use only.

Waste Management

Refer to local legal requirements.

Preparation

The sample dilution matrix is ready to use.

Mix 1 part of human sera + 20 parts of sample dilution matrix (e.g. 10 µL sample material + 200 µL sample dilution matrix).

Materials Required

General laboratory equipment



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com