

## Cystatin C FS\*

### Order Information

<b>Cat. No.</b>	<b>Kit size</b>
1 7158 99 10 966	200 (R1: 2 x 100, R2: 2 x 100)

### Intended Use

Diagnostic reagent for quantitative in vitro determination of cystatin C in human serum or heparin plasma on automated BioMajesty® JCA-BM6010/C.

### Summary

Cystatin C is a non-glycosylated, basic protein with a low molecular weight of 13 kDa. It acts as a cysteine protease inhibitor, is endogenously produced at a constant rate by all nucleated cells investigated and freely filtered by the glomerular membrane before being almost completely reabsorbed and degraded in the renal tubuli. Cystatin C is suggested to be a better marker for detection of reduced glomerular filtration rate (GFR) than creatinine especially for the detection of a moderate impairment of kidney function. The cystatin C blood level is, in contrast to creatinine, is less dependent on factors such as sex, muscle mass and age. Cystatin C determination may be useful especially in children, elder people, in diabetics, in patients with liver cirrhosis, in adult renal transplant recipients, in cancer patients and in pregnant woman suspected of preeclampsia. [1-9]

### Method

Particle enhanced immunoturbidimetric test

Determination of cystatin C concentration by photometric measurement of antigen antibody reaction between antibodies against cystatin C bound to polystyrene particles and cystatin C present in the sample.

### Reagents

#### Components and Concentrations

<b>R1:</b>	TRIS	pH 7.5	100 mmol/L
	NaCl		200 mmol/L
<b>R2:</b>	Borate		7.5 mmol/L
	Monoclonal antibodies (mouse)		< 1%
	against human cystatin C bound to		
	carboxylated polystyrene particles		

### Storage and Stability

Reagents are stable up to the date of expiry indicated on the kit, if stored at 2 – 8°C and contamination is avoided. Protect from light.

### Warnings and Precautions

1. Reagent 1 contains sodium azide (0.9 g/L) as preservative. Do not swallow! Avoid contact with skin and mucous membranes.
2. Reagent 2 contains sodium azide (0.95 g/L) as preservative. Do not swallow! Avoid contact with skin and mucous membranes.
3. The reagents contain animal material. Handle the product as potentially infectious according to universal precautions and good clinical laboratory practice.
4. In very rare cases, samples of patients with gammopathy might give falsified results [10].
5. Please refer to the safety data sheets and take the necessary precautions for the use of laboratory reagents. For diagnostic purposes, the results should always be assessed with the patient's medical history, clinical examinations and other findings.
6. For professional use only.

### Waste Management

Refer to local legal requirements.

### Reagent Preparation

The reagents are ready to use. The bottles are placed directly into the reagent rotor.

### Materials Required

General laboratory equipment

### Specimen

Human serum or heparin plasma

Stability [11]:

2 days	at	20 – 25°C
1 week	at	2 – 8°C
1 month	at	-20°C

Only freeze once. Discard contaminated specimens.

### Calibrators and Controls

DiaSys TruCal Cystatin C calibrator set is recommended for calibration. Calibrator values have been made traceable to the IFCC reference material ERM®-DA471. Use DiaSys TruLab Cystatin C Level 1 and Level 2 for internal quality control. Each laboratory should establish corrective action in case of deviations in control recovery.

	Cat. No.	Kit size
TruCal Cystatin C	1 7150 99 10 059	5 x 1 mL
TruLab Cystatin C Level 1	5 9870 99 10 046	3 x 1 mL
TruLab Cystatin C Level 2	5 9880 99 10 046	3 x 1 mL

### Performance Characteristics

Exemplary data mentioned below may slightly differ in case of deviating measurement conditions.

Measuring range up to 8 mg/L, depending on the concentration of the highest calibrator. In case of higher concentrations re-measure samples after manual dilution with NaCl solution (9 g/L) or use rerun function.		
Limit of detection**	0.2 mg/L	
No prozone effect up to 30 mg/L.		
Onboard stability	12 weeks	
Calibration stability	12 weeks	
Interfering substance	Interferences ≤ 10% up to	Analyte concentration [mg/L]
Bilirubin (conjugated)	50 mg/dL	0.451
	50 mg/dL	4.61
Bilirubin (unconjugated)	50 mg/dL	0.471
	50 mg/dL	4.86
Hemoglobin	600 mg/dL	0.557
	1000 mg/dL	4.43
Lipemia (triglycerides)	1200 mg/dL	0.652
	2000 mg/dL	6.09
Rheumatoid factor	600 IU/mL	0.499
	600 IU/mL	4.56
Thyroid dysfunction impacts cystatin C levels [12].		
For further information on interfering substances refer to Young DS [13,14].		

Precision (Serum)			
Within run (n=20)	Sample 1	Sample 2	Sample 3
Mean [mg/L]	0.554	0.896	4.33
CV [%]	2.15	1.06	0.505
Total precision CLSI (n=80)	Sample 1	Sample 2	Sample 3
Mean [mg/L]	0.581	0.938	4.54
CV [%]	2.70	1.71	1.09

Method comparison (Serum; n=99)	
Test x	Competitor Cystatin C (BN ProSpec®)
Test y	DiaSys Cystatin C FS (BioMajesty® JCA-BM6010/C)
Slope	0.982
Intercept	-0.001 mg/L
Coefficient of correlation	0.998

\*\* lowest measurable concentration which can be distinguished from zero; mean + 3 SD (n = 20) of an analyte free specimen.

## Reference Range [15,16,17]

	[mg/L]
<b>Children</b>	
4th and 5th day	1.22 – 1.68
< 1 month	1.37 – 1.89
1 – 12 months	0.73 – 1.17
> 12 months	0.60 – 0.84

Mean values +/- 1 SD (standard deviation) are listed

### Adults

19 – 49 years	0.53 – 0.92
≥ 50 years	0.58 – 1.02

Indication of 2 SD range

Each laboratory should check if the reference ranges are transferable to its own patient population and determine own reference ranges if necessary.

## Literature

1. Erland J.E., Randers E., Kristensen J.H. Reference intervals for serum cystatin C and serum creatinine in adults. *Clin Chem Lab Med* 1998; 36(6):393-397.
2. Lamb E., Newman DJ, Price CP Kidney function tests. In: Burtis CA, Ashwood ER, Bruns DE, editors. *Tietz Textbook of Clinical Chemistry and Molecular Diagnostics*. 4th edition St. Louis Missouri: Elsevier Saunders; 2006; p. 823-835.
3. Kyhse-Andersen, Schmidt C., Nordin G. et al. Serum cystatin C, determined by a rapid, automated particle-enhanced turbidimetric method, is a better marker than serum creatinine for glomerular filtration rate. *ClinChem* 1994; 40(10):1921-6.
4. Le Bricon T., Leblanc I et al. Evaluation of renal function in intensive care: plasma cystatin C vs. creatinine and derived glomerular filtration rates *Clin Chem Lab Med* 2005; 43(9):953-957.
5. Le Bricon T., Thervet E., Benlakehal M. et al. Changes in Plasma cystatin C after renal transplantation and acute rejection in adults. *Clin Chem* 1999; 45(12):2243-9
6. Ustundag Y., Samsar U. et al. Analysis of glomerular filtration rate, serum cystatin C levels, and resistive index values in cirrhosis patients. *Clin Chem Lab Med* 2007; 45(7):890-94.
7. Stabuc B., Vrhovec L. et al. Improved prediction of decreased creatinine clearance by serum cystatin C : Use in cancer patients before and during chemotherapy. *Clin Chem* 2000; 46(2):193-7.
8. Pucci L., Triscornia S., Lucchesi D. et al. Cystatin C and estimates of renal function: searching for a better measure of kidney function in diabetic patients. *Clin Chem* 2007; 53(3):480-8.
9. Strevens H, Wide-Svensson D, Grubb A. Serum cystatin C is a better marker for preeclampsia than serum creatinine or serum urate. *Scand J Clin Lab Invest* 2001; 61:575-80.
10. Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: mechanisms, detection and prevention. *ClinChemLabMed* 2007;45(9):1240-1243.
11. Guder WG, Zawta B. et al. *The Quality of Diagnostic Samples*. 1st ed. Darmstadt: GIT Verlag; 2001; p. 24-5.
12. Wiesli P., Schwegler B. et al. Serum cystatin C is sensitive to small changes in thyroid function. *Clinica Chimica Acta* 2003; 338: 87-90.
13. Young DS. *Effects of Drugs on Clinical Laboratory Tests*. 5th ed. Volume 1 and 2. Washington, DC: The American Association for Clinical Chemistry Press 2000.
14. Young DS. *Effects on Clinical Laboratory Tests - Drugs Disease, Herbs & Natural Products*, <https://clinfx.wiley.com/aaccweb/aacc/>, accessed in August 2021. Published by AACC Press and John Wiley and Sons, Inc.
15. Bökenkamp A, Grabensee A, Stoffel-Wagner B, Hasan C, Henne T, Offner G, Lentze MJ. The  $\beta$ 2-microglobulin/cystatin

- C ratio: a potential marker of post-transplant lymphoproliferative disease. *Clin Nephrol* 2001; 58: 417-22.
16. Randers E, Krue S, Erlandsen EJ, Danielsen H, Hansen H, Hansen LG. Reference interval for serum cystatin C in children. *ClinChem* 1999; 45:1856-60.
  17. Finney H, Newman DJ, Price CP. Adult reference ranges for serum cystatin C, creatinine and predicted creatinine clearance. *Ann Clin Biochem* 2000; 37: 49-59.



DiaSys Diagnostic Systems GmbH  
 Alte Strasse 9 65558 Holzheim  
 Germany  
[www.diasys-diagnostics.com](http://www.diasys-diagnostics.com)

\* Fluid Stable

## Cystatin C FS

Chemistry code 10 715

### Application for serum samples

This application was set up and evaluated by DiaSys. It is based on the standard equipment at that time and does not apply to any equipment modifications undertaken by unqualified personnel.

Analytical Conditions	
R1 volume	90
R2e volume	0
R2 volume	30
R1 diluent vol	0
R2e diluent vol	0
R2 diluent vol	0
Sample vol (S)	1.0
Sample vol (U)	1.0
Reagent 1 mix	strong
Reagent 2e mix	weak
Reagent 2 mix	strong
Reaction time	10

Endpoint Method	
Re.absorb (u)	9.999
Re.absorb (d)	-9.999

Calculation Method Setting	
M-DET.P.l	0
M-DET.P.m	41
M-DET.P.n	42
S-DET.P.p	22
S-DET.P.r	23
Check D.P.l.	0
Limit value	0.003
Variance	10
Reac.type	Inc

Sub-analy. Conditions	
Name	CYSC
Digits	3
M-wave L.	596
S-wave.L	****
Analy.mthd.	EPA
Calc.mthd.	MSTD
Qualit. judge	No

Reaction Rate Method	
Cycle	2
Factor	2
E2 corre	Not do
Blank (u)	9.999
Blank (d)	-9.999
Sample (u)	9.999
Sample (d)	-9.999

Analysis Test Condition Setting (M)		
Sample Type	Serum	Urine
Reac. sample vol.	1.0	1.0
Diluent method	No dil	No dil
Undil. sample vol.	0	0
Diluent volume	0	0
Diluent position	0	0

Prozone	
Prozone form	No
Prozone limit	9.999
Prozone judge	Upper limit
Judge limit	9.999
M-DET.P.m	0
M-DET.P.n	0
S-DET.P.p	0
S-DET.P.r	0

MULTI-STD Setting								
Formula	Spline	Axis Conv	No conv					
Blank	Blank-any value	Points	6					
	FV	Reac. smp. vol.	Dil. method	Dil. smp. vol.	Diluent vol.	Diluent pos.	STD H	STD L
BLK	#	1.0	No dil	0	0	0	9.999	-9.999
1	#	1.0	No dil	0	0	0	9.999	-9.999
2	#	1.0	No dil	0	0	0	9.999	-9.999
3	#	1.0	No dil	0	0	0	9.999	-9.999
4	#	1.0	No dil	0	0	0	9.999	-9.999
5	#	1.0	No dil	0	0	0	9.999	-9.999

# entered by user