

# Immunoglobulin E FS\*

## Order Information

Cat. No.	Kit size
1 7239 99 10 921	320 (4 x 80)
1 7239 99 10 926	160 (2 x 80)

## Intended Use

Diagnostic reagent for quantitative in vitro determination of immunoglobulin E (IgE) in human serum or heparin plasma on automated DiaSys respons<sup>®</sup>910.

## Summary

The human immunoglobulin classes (IgG, IgA, IgM, IgE and IgD) are a group of functionally and structurally closely related glycoproteins. Human IgE has a molecular weight of about 190 000 dalton and consists of two identical heavy chains and two identical light chains which are bound together by disulfide bonds in a characteristic Y-shaped form. The original function of IgE is the specific defense of parasites. In the developed countries, it plays a major role in the mediation of immediate type hypersensitivity reactions (type I according to Coombs and Gell). Harmless, polyvalent antigens (pollen, house dust mites), stimulate B cells at the site of entry to synthesize specific IgE which in part binds to mast cells. The half life of unbound IgE is 2 – 3 days while mast cell-bound IgE has a half-life from months to years. During the next contact of the antigen with the sensitized mast cell, bound IgE are cross-linked. The cell is degranulated and mediators (mainly histamine) are released which cause, for example, symptoms of hay fever, asthma, and atopic eczema. Elevated IgE levels occur in atopic diseases, parasitic infection, diseases with T cell dysfunction (e.g. AIDS), certain malignant tumors (respiratory tract, gastrointestinal tract), hyper-IgE syndrome, graft-versus-host disease, and in severe burns. Measurement of total IgE is mainly conducted to diagnose of atopic diseases where highly increased IgE levels may occur. IgE testing is a good tool especially in differential diagnostic examination of clinical pictures with possible allergic background [1].

## Method

Particle enhanced immunoturbidimetric test

Determination of IgE concentration by photometric measurement of antigen antibody reaction of latex particles coated with antibodies to human IgE with IgE present in the sample.

## Reagents

### Components and Concentrations

<b>R1:</b>	Glycine	pH 8.3	170 mmol/L
	NaCl		100 mmol/L
<b>R2:</b>	Glycine	pH 7.3	170 mmol/L
	NaCl		100 mmol/L
	Latex particles coated with anti-human IgE monoclonal antibody (mouse)		1.3 g/L

## Storage and Stability

Reagents are stable up to the date of expiry indicated on the kit, if stored at 2 – 8°C and contamination is avoided. Do not freeze and protect from light.

## Warnings and Precautions

- The reagents contain sodium azide (< 0.1%) as preservative. Do not swallow! Avoid contact with skin and mucous membranes.
- The reagents contain animal material. Handle the product as potentially infectious according to universal precautions and good clinical laboratory practice.
- Heterophile antibodies in patient samples may cause falsified results.
- In very rare cases, samples of patients with gammopathy might give falsified results [2].
- Please refer to the safety data sheets and take the necessary precautions for the use of laboratory reagents. For diagnostic purposes, the results should always be assessed with the patient's medical history, clinical examinations and other findings.
- For professional use only.

## Waste Management

Refer to local legal requirements.

## Reagent Preparation

The reagents are ready to use. The bottles are placed directly into the reagent rotor.

## Materials Required

General laboratory equipment

## Specimen

Human serum or heparin plasma

Stability [3]:

7 days	at	20 – 25°C
7 days	at	4 – 8°C
6 months	at	-20°C

Only freeze once. Discard contaminated specimens.

## Calibrators and Controls

DiaSys TruCal IgE calibrator set is recommended for calibration. Calibrator values have been made traceable to the WHO Reference Material NIBSC 75/502. Use DiaSys TruLab Protein Level 1 and Level 2 for internal quality control. Each laboratory should establish corrective action in case of deviations in control recovery.

	Cat. No.	Kit size
TruCal IgE	1 7230 99 10 059	5 x 1 mL
TruLab Protein Level 1	5 9500 99 10 046	3 x 1 mL
TruLab Protein Level 2	5 9510 99 10 046	3 x 1 mL

## Performance Characteristics

Exemplary data mentioned below may slightly differ in case of deviating measurement conditions.

Measuring range up to 1000 IU/mL, depending on the concentration of the highest calibrator.		
In case of higher concentrations re-measure samples after manual dilution with NaCl solution (9 g/L) or use rerun function.		
Limit of detection**	16 IU/mL	
No prozone effect up to 15000 IU/mL.		
Onboard stability	10 days	
Calibration stability	10 days	
Interfering substance	Interferences ≤ 10% up to	Analyte concentration [IU/mL]
<b>Bilirubin</b> (conjugated)	60 mg/dL	70.4
	60 mg/dL	195
<b>Bilirubin</b> (unconjugated)	60 mg/dL	71.1
	60 mg/dL	203
<b>Hemoglobin</b>	1200 mg/dL	70.2
	1200 mg/dL	193
<b>Lipemia</b> (triglycerides)	1000 mg/dL	140
	800 mg/dL	202
For further information on interfering substances refer to Young DS [4,5].		

Precision			
Within run (n=20)	Sample 1	Sample 2	Sample 3
Mean [IU/mL]	79.5	228	684
CV [%]	2.56	1.36	0.89
Between day (n=20)	Sample 1	Sample 2	Sample 3
Mean [IU/mL]	84.1	231	531
CV [%]	4.60	1.92	2.11

Method comparison (n=103)	
Test x	DiaSys Immunoglobulin E FS (Hitachi 917)
Test y	DiaSys Immunoglobulin E FS (respons <sup>®</sup> 910)
Slope	1.03
Intercept	-1.46 IU/mL
Coefficient of correlation	0.999

\*\* according to CLSI document EP17-A, Vol. 24, No. 34

## Reference Range [6,7]

Age group	Upper limit of the normal range (95th percentile)
Newborns	1.5 IU/mL
1st year	15 IU/mL
1 – 5 years	60 IU/mL
6 – 9 years	90 IU/mL
10 – 15 years	200 IU/mL
Adults	100 IU/mL

Each laboratory should check if the reference ranges are transferable to its own patient population and determine own reference ranges if necessary.

## Literature

1. Thomas L. Clinical Laboratory Diagnostics. 1st ed. Frankfurt: TH-Books Verlagsgesellschaft; 1998. p. 667-78,774-85.
2. Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: Mechanism, detection and prevention. Clin Chem Lab Med 2007; 45(9): 1240-1243.
3. Guder WG, Zawta B et al. The Quality of Diagnostic Samples. 1st ed. Darmstadt: GIT Verlag; 2001. p. 34-5.
4. Young DS. Effects of Drugs on Clinical Laboratory Tests. 5th ed. Volume 1 and 2. Washington, DC: The American Association for Clinical Chemistry Press 2000.
5. Young DS. Effects on Clinical Laboratory Tests - Drugs Disease, Herbs & Natural Products, <https://clinf.wiley.com/aaccweb/aacc/>, accessed in September 2021. Published by AACC Press and John Wiley and Sons, Inc.
6. Ringel KP, Dati F, Buchholz E. IgE-Normalwerte bei Kindern, Laboratoriumsblätter 1982;32:26-34.
7. Dati F, Ringel KP. Reference values for serum IgE in healthy non-atopic children and adults. Clin Chem 1982; 28:1556.



DiaSys Diagnostic Systems GmbH  
 Alte Strasse 9 65558 Holzheim  
 Germany  
[www.diasys-diagnostics.com](http://www.diasys-diagnostics.com)

\* Fluid Stable

## Immunoglobulin E FS

### Application for serum and plasma samples

This application was set up and evaluated by DiaSys. It is based on the standard equipment at that time and does not apply to any equipment modifications undertaken by unqualified personnel.

Identification	
This method is usable for analysis:	Yes
Twin reaction:	No
Name:	IGE
Shortcut:	
Reagent barcode reference:	717
Host reference:	

Technic	
Type:	Fixed time kinetic
First reagent:[ $\mu$ L]	160
Blank reagent	Yes
Sensitive to light	
Second reagent:[ $\mu$ L]	80
Blank reagent	Yes
Sensitive to light	
Main wavelength:[nm]	600
Secondary wavelength:[nm]	
Polychromatic factor:	
1 st reading time [min:sec]	(04:48)
Last reading time [min:sec]	07:24
Reaction way:	Increasing
Linear Kinetics	
Substrate depletion: Absorbance li	
Linearity: Maximum deviation [%]	
Fixed Time Kinetics	
Substrate depletion: Absorbance limit	
Endpoint	
Stability: Largest remaining slope	
Prozone Limit [%]	

Reagents	
Decimals	
Units	

Sample	
Diluent	DIL A (NaCl)
Hemolysis:	
Agent [ $\mu$ L]	0 (no hemolysis)
Cleaner	
Sample [ $\mu$ L]	0
Technical limits	
Concentration technical limits-Lower	16
Concentration technical limits-Upper	1000
SERUM	
Normal volume [ $\mu$ L]	8
Normal dilution (factor)	1
Below normal volume [ $\mu$ L]	16
Below normal dilution (factor)	1
Above normal volume [ $\mu$ L]	8
Above normal dilution (factor)	6
URIN	
Normal volume [ $\mu$ L]	8
Normal dilution (factor)	1
Below normal volume [ $\mu$ L]	16
Below normal dilution (factor)	1
Above normal volume [ $\mu$ L]	8
Above normal dilution (factor)	6
PLASMA	
Normal volume [ $\mu$ L]	8
Normal dilution (factor)	1
Below normal volume [ $\mu$ L]	16
Below normal dilution (factor)	1
Above normal volume [ $\mu$ L]	8
Above normal dilution (factor)	6
CSF	
Normal volume [ $\mu$ L]	8
Normal dilution (factor)	1
Below normal volume [ $\mu$ L]	16
Below normal dilution (factor)	1
Above normal volume [ $\mu$ L]	8
Above normal dilution (factor)	6
Whole blood	
Normal volume [ $\mu$ L]	8
Normal dilution (factor)	1
Below normal volume [ $\mu$ L]	16
Below normal dilution (factor)	1
Above normal volume [ $\mu$ L]	8
Above normal dilution (factor)	6

Results	
Decimals	1
Units	IU/mL
Correlation factor-Offset	0.000
Correlation factor-Slope	1.000

Range	
Gender	All
Age	Adults
SERUM	>= <=100
URINE	
PLASMA	>= <=100
CSF	
Whole blood	
Gender	
Age	
SERUM	
URINE	
PLASMA	
CSF	
Whole blood	

Contaminants	
Please refer to r910 Carryover Pair Table	

Calibrators details		
Calibrator list	Concentration	
Cal. 1/Blank	0	
Cal. 2	*	
Cal. 3	*	
Cal. 4	*	
Cal. 5	*	
Cal. 6	*	
	Max delta abs.	
Cal. 1	0.0100	
Cal. 2	0.0100	
Cal. 3	0.0100	
Cal. 4	0.0100	
Cal. 5	0.0150	
Cal. 6	0.0250	
Drift limit [%]	7.0	

Calculations	
Model	Logit-Log 5P
Degree	Auto

\* Enter calibrator value