

TruLab L Level 2

LOT 38576

Cat. No. 5 9030 99 10 065 (3 x 3 mL)



2028-08-31



Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 1

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
APOA1 Apolipoprotein A1	Immunoturbidimetric	1 7102	198 1.98	158 - 237 1.58 - 2.37	mg/dL g/L
APOB Apolipoprotein B	Immunoturbidimetric	1 7112	135 1.35	108 - 162 1.08 - 1.62	mg/dL g/L
CHOL Cholesterol	CHOD-PAP	1 1300	240 6.21 2.40	206 - 274 5.34 - 7.08 2.06 - 2.74	mg/dL mmol/L g/L
HDLCD HDL-Cholesterol	Direct, colorimetric	1 3561	55.7 1.44 0.557	44.6 - 66.8 1.15 - 1.73 0.446 - 0.668	mg/dL mmol/L g/L
	CHOD-PAP Precipitant	1 1300 + 1 3540	56.4 1.46 0.564	45.1 - 67.6 1.17 - 1.75 0.451 - 0.676	mg/dL mmol/L g/L
LDLCD LDL-Cholesterol	Direct, colorimetric	1 4131	148 3.82 1.48	118 - 177 3.05 - 4.58 1.18 - 1.77	mg/dL mmol/L g/L
	CHOD-PAP Precipitant	1 1300 + 1 4330	152 3.93 1.52	122 - 182 3.15 - 4.72 1.22 - 1.82	mg/dL mmol/L g/L
Lp-PLA₂ Lipoprotein-associated phospholipase A2	Colorimetric reconstitution volume.: 1.0 mL	1 7181	1088 18.1	870 - 1305 14.5 - 21.8	U/L µkat/L
	Colorimetric reconstitution volume.: 1.5 mL	1 7181	829 13.8	663 - 994 11.05 - 16.6	U/L µkat/L
NEFA Non-esterified fatty acids	Enzymatic	1 5781	50.0 1.77 0.500	40.0 - 60.0 1.41 - 2.12 0.400 - 0.600	mg/dL mmol/L g/L
TRIG Triglycerides	GPO-PAP	1 5710	253 2.85 2.53	202 - 303 2.28 - 3.42 2.02 - 3.03	mg/dL mmol/L g/L



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab L Level 2

LOT 38576

Cat. No. 5 9030 99 10 065 (3 x 3 mL)



2028-08-31

respons[®]910



Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 1

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
APOA1 Apolipoprotein A1	Immunoturbidimetric	1 7102	198 1.98	158 - 237 1.58 - 2.37	mg/dL g/L
APOB Apolipoprotein B	Immunoturbidimetric	1 7112	135 1.35	108 - 162 1.08 - 1.62	mg/dL g/L
CHOL Cholesterol	CHOD-PAP	1 1300	240 6.21 2.40	206 - 274 5.34 - 7.08 2.06 - 2.74	mg/dL mmol/L g/L
HDLCD HDL-Cholesterol	Direct, colorimetric	1 3561	55.7 1.44 0.557	44.6 - 66.8 1.15 - 1.73 0.446 - 0.668	mg/dL mmol/L g/L
LDLCD LDL-Cholesterol	Direct, colorimetric	1 4131	148 3.82 1.48	118 - 177 3.05 - 4.58 1.18 - 1.77	mg/dL mmol/L g/L
Lp-PLA₂ Lipoprotein-associated phospholipase A2	Colorimetric reconstitution volume.: 1.0 mL	1 7181	1088 18.1	870 - 1305 14.5 - 21.8	U/L µkat/L
	Colorimetric reconstitution volume.: 1.5 mL	1 7181	829 13.8	663 - 994 11.05 - 16.6	U/L µkat/L
NEFA Non-esterified fatty acids	Enzymatic	1 5781	50.0 1.77 0.500	40.0 - 60.0 1.41 - 2.12 0.400 - 0.600	mg/dL mmol/L g/L
TRIG Triglycerides	GPO-PAP	1 5710	253 2.85 2.53	202 - 303 2.28 - 3.42 2.02 - 3.03	mg/dL mmol/L g/L




DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab L Level 2

LOT 38576

Cat. No. 5 9030 99 10 065 (3 x 3 mL)

 2028-08-31

respons[®]920



Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 1

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	1SD	Unit
APOA1 Apolipoprotein A1	Immunoturbidimetric	1 7102	198 1.98	158 - 237 1.58 - 2.37	13.2 0.132	mg/dL g/L
APOB Apolipoprotein B	Immunoturbidimetric	1 7112	135 1.35	108 - 162 1.08 - 1.62	9.02 0.090	mg/dL g/L
CHOL Cholesterol	CHOD-PAP	1 1300	240 6.21 2.40	206 - 274 5.34 - 7.08 2.06 - 2.74	11.2 0.290 0.112	mg/dL mmol/L g/L
HDLCD HDL-Cholesterol	Direct, colorimetric	1 3561	55.7 1.44 0.557	44.6 - 66.8 1.15 - 1.73 0.446 - 0.668	3.71 0.096 0.037	mg/dL mmol/L g/L
LDLCD LDL-Cholesterol	Direct, colorimetric	1 4131	148 3.82 1.48	118 - 177 3.05 - 4.58 1.18 - 1.77	9.84 0.254 0.098	mg/dL mmol/L g/L
TRIG Triglycerides	GPO-PAP	1 5710	253 2.85 2.53	202 - 303 2.28 - 3.42 2.02 - 3.03	16.9 0.190 0.169	mg/dL mmol/L g/L




DiaSys Diagnostic Systems GmbH
 Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab L Level 2

LOT 38576

Cat. No. 5 9030 99 10 065 (3 x 3 mL)

 2028-08-31

respons[®]940



Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 1

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	1SD	Unit
APOA1 Apolipoprotein A1	Immunoturbidimetric	1 7102	198 1.98	158 - 237 1.58 - 2.37	13.2 0.132	mg/dL g/L
APOB Apolipoprotein B	Immunoturbidimetric	1 7112	135 1.35	108 - 162 1.08 - 1.62	9.02 0.090	mg/dL g/L
CHOL Cholesterol	CHOD-PAP	1 1300	240 6.21 2.40	206 - 274 5.34 - 7.08 2.06 - 2.74	11.2 0.290 0.112	mg/dL mmol/L g/L
HDLCD HDL-Cholesterol	Direct, colorimetric	1 3561	55.7 1.44 0.557	44.6 - 66.8 1.15 - 1.73 0.446 - 0.668	3.71 0.096 0.037	mg/dL mmol/L g/L
LDLCD LDL-Cholesterol	Direct, colorimetric	1 4131	148 3.82 1.48	118 - 177 3.05 - 4.58 1.18 - 1.77	9.84 0.254 0.098	mg/dL mmol/L g/L
TRIG Triglycerides	GPO-PAP	1 5710	253 2.85 2.53	202 - 303 2.28 - 3.42 2.02 - 3.03	16.9 0.190 0.169	mg/dL mmol/L g/L




DiaSys Diagnostic Systems GmbH
 Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab L Level 2

LOT 38576

Cat. No. 5 9030 99 10 065 (3 x 3 mL)

 2028-08-31

BioMajesty[®]

  **IVD**

Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 1

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	1SD	Unit
APOA1 Apolipoprotein A1	Immunoturbidimetric	1 7102	198 1.98	158 - 237 1.58 - 2.37	13.2 0.132	mg/dL g/L
APOB Apolipoprotein B	Immunoturbidimetric	1 7112	135 1.35	108 - 162 1.08 - 1.62	9.02 0.090	mg/dL g/L
CHOL Cholesterol	CHOD-PAP	1 1300	240 6.21 2.40	206 - 274 5.34 - 7.08 2.06 - 2.74	11.2 0.290 0.112	mg/dL mmol/L g/L
HDLCD HDL-Cholesterol	Direct, colorimetric	1 3561	55.7 1.44 0.557	44.6 - 66.8 1.15 - 1.73 0.446 - 0.668	3.71 0.096 0.037	mg/dL mmol/L g/L
LDLCD LDL-Cholesterol	Direct, colorimetric	1 4131	148 3.82 1.48	118 - 177 3.05 - 4.58 1.18 - 1.77	9.84 0.254 0.098	mg/dL mmol/L g/L
NEFA Non-esterified fatty acids	Enzymatic	1 5781	50.0 1.77 0.500	40.0 - 60.0 1.41 - 2.12 0.400 - 0.600	3.33 0.118 0.033	mg/dL mmol/L g/L
TRIG Triglycerides	GPO-PAP	1 5710	253 2.85 2.53	202 - 303 2.28 - 3.42 2.02 - 3.03	16.9 0.190 0.169	mg/dL mmol/L g/L



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com